

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
 Data File : VX040128.D
 Acq On : 05 Feb 2024 19:54
 Operator : JC/MD
 Sample : P1355-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-214

Quant Time: Feb 06 03:06:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	90299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	157129	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	142253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63210	43.496	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.000%
35) Dibromofluoromethane	5.385	113	51993	48.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.220%
50) Toluene-d8	8.647	98	205125	51.280	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.560%
62) 4-Bromofluorobenzene	11.079	95	77762	50.486	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.980%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	12397	18.206	ug/l	92
20) Methylene Chloride	2.788	84	4726	3.906	ug/l	96
43) Isopropyl Acetate	6.342	43	68449	25.692	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040128.D
Acq On : 05 Feb 2024 19:54
Operator : JC/MD
Sample : P1355-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-214

Quant Time: Feb 06 03:06:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

