

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
 Data File : VX040102.D
 Acq On : 05 Feb 2024 09:54
 Operator : JC/MD
 Sample : VX0205MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0205MBL01

Quant Time: Feb 06 02:54:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	90819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162224	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	150903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73247	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66353	45.397	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.800%
35) Dibromofluoromethane	5.385	113	53400	48.356	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.720%
50) Toluene-d8	8.647	98	192344	46.574	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.140%
62) 4-Bromofluorobenzene	11.079	95	72678	45.704	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.400%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040102.D
Acq On : 05 Feb 2024 09:54
Operator : JC/MD
Sample : VX0205MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0205MBL01

Quant Time: Feb 06 02:54:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 712
Delta R.T. -0.000 min
Lab File: VX040102.D
Acq: 05 Feb 2024 09:54




Instrument :
MSVOA_X
ClientSampleId :
VX0205MBL01


Tgt Ion:168 Resp: 90819
Ion Ratio Lower Upper
168 100
99 61.3 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 45.397 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX040102.D
Acq: 05 Feb 2024 09:54

Tgt Ion: 65 Resp: 66353
Ion Ratio Lower Upper
65 100
67 52.3 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 911
Delta R.T. -0.006 min
Lab File: VX040102.D
Acq: 05 Feb 2024 09:54




Instrument :
MSVOA_X
ClientSampleId :
VX0205MBL01


Tgt Ion:114 Resp: 162224
Ion Ratio Lower Upper
114 100
63 21.3 0.0 45.0
88 16.3 0.0 36.6



#35
Dibromofluoromethane
Concen: 48.356 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX040102.D
Acq: 05 Feb 2024 09:54

Tgt Ion:113 Resp: 53400
Ion Ratio Lower Upper
113 100
111 102.5 82.9 124.3
192 16.9 13.3 19.9









#50
Toluene-d8
Concen: 46.574 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX040102.D
Acq: 05 Feb 2024 09:54




Instrument :
MSVOA_X
ClientSampleId :
VX0205MBL01


Tgt Ion: 98 Resp: 192344
Ion Ratio Lower Upper
98 100
100 65.2 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 45.704 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX040102.D
Acq: 05 Feb 2024 09:54

Tgt Ion: 95 Resp: 72678
Ion Ratio Lower Upper
95 100
174 73.5 0.0 135.6
176 70.1 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX040102.D
Acq: 05 Feb 2024 09:54

Instrument :
MSVOA_X
ClientSampleId :
VX0205MBL01

Tgt Ion:117 Resp: 150903
Ion Ratio Lower Upper
117 100
82 57.5 46.2 69.4
119 31.5 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX040102.D
Acq: 05 Feb 2024 09:54

Tgt Ion:152 Resp: 73247
Ion Ratio Lower Upper
152 100
115 62.7 43.3 129.9
150 155.6 0.0 345.0

