

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX020918\
 Data File : VX000050.D
 Acq On : 09 Feb 2018 18:56
 Operator : JC/MD
 Sample : VX0209WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ClientSampleId :
 VX0209WBL01

Quant Time: Feb 10 00:07:22 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X020918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 09 23:01:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	131214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	212895	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	196998	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	104341	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	70872	51.69	ug/l	0.00
Spiked Amount			Recovery	=	103.38%	
35) Dibromofluoromethane	5.51	113	60913	49.00	ug/l	0.00
Spiked Amount			Recovery	=	98.00%	
50) Toluene-d8	8.73	98	219002	47.34	ug/l	0.00
Spiked Amount			Recovery	=	94.68%	
62) 4-Bromofluorobenzene	11.15	95	86370	45.91	ug/l	0.00
Spiked Amount			Recovery	=	91.82%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	1.66	94	405	0.56	ug/l	82
6) Chloroethane	1.73	64	903	1.10	ug/l #	83
10) Methyl Iodide	2.52	142	831	2.19	ug/l #	85
13) Acrolein	2.29	56	131	0.47	ug/l #	58
15) Acrylonitrile	3.15	53	282	0.30	ug/l #	91
16) Acetone	2.45	43	571	0.70	ug/l #	87
17) Carbon Disulfide	2.57	76	2012	0.54	ug/l #	86
21) trans-1,2-Dichloroethene	3.17	96	321	0.23	ug/l #	48
36) 1,1-Dichloropropene	5.81	75	423	0.23	ug/l #	50
38) Carbon Tetrachloride	5.68	117	12715	6.11	ug/l #	16
39) Methylcyclohexane	7.47	83	452	0.20	ug/l	92
60) Dibromochloromethane	9.34	129	394	0.23	ug/l #	11
64) Tetrachloroethene	9.34	164	402	0.27	ug/l	86
68) m/p-Xylenes	10.37	106	771	0.29	ug/l	99
76) 1,2,3-Trichloropropane	11.15	75	44194	22.49	ug/l #	31
78) n-propylbenzene	11.37	91	1652	0.22	ug/l	94
81) trans-1,4-Dichloro-2-buten	11.15	75	44194	55.65	ug/l #	11
82) 4-Chlorotoluene	11.52	91	1377	0.26	ug/l	89
85) sec-Butylbenzene	11.96	105	1400	0.20	ug/l	99
86) p-Isopropyltoluene	12.07	119	1531	0.25	ug/l #	74
87) 1,3-Dichlorobenzene	12.03	146	873	0.26	ug/l	80
88) 1,4-Dichlorobenzene	12.03	146	873	0.25	ug/l	80
89) n-Butylbenzene	12.40	91	2118	0.37	ug/l	97
93) 1,2,4-Trichlorobenzene	13.66	180	1100	0.43	ug/l	96
94) Hexachlorobutadiene	13.79	225	463	0.39	ug/l	96
96) 1,2,3-Trichlorobenzene	14.03	180	657	0.27	ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX020918\
Data File : VX000050.D
Acq On : 09 Feb 2018 18:56
Operator : JC/MD
Sample : VX0209WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ClientSampleId :
VX0209WBL01

Quant Time: Feb 10 00:07:22 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X020918W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 09 23:01:05 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.68 min Scan# 835

Delta R.T. -0.00 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56


Instrument :
ClientSampleId :
VX0209WBL01

Tgt Ion: 168 Resp: 131214

Ion Ratio Lower Upper




168 100

99 51.7 39.4 59.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#5

Bromomethane

Concen: 0.56 ug/l

RT: 1.66 min Scan# 176

Delta R.T. 0.02 min

Lab File: VX000050.D


Acq: 09 Feb 2018 18:56

Tgt Ion: 94 Resp: 405

Ion Ratio Lower Upper



94 100


96 113.0 76.7 115.1



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0





#6

Chloroethane

Concen: 1.10 ug/l

RT: 1.73 min Scan# 187

Delta R.T. 0.01 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56


Instrument :
ClientSampleId :
VX0209WBL01

Tgt Ion: 64 Resp: 903




Ion Ratio Lower Upper

64 100

66 23.6 26.4 39.6#



Abundance Ion 63.90 (63.60 to 64.60): VX0
Ion 65.90 (65.60 to 66.60): VX0



#10

Methyl Iodide

Concen: 2.19 ug/l

RT: 2.52 min Scan# 317

Delta R.T. 0.01 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56


Tgt Ion: 142 Resp: 831

Ion Ratio Lower Upper



142 100


127 39.0 36.8 55.2

141 3.0 11.4 17.0#



Abundance Ion 141.80 (141.50 to 142.50): V
Ion 126.80 (126.50 to 127.50): V
Ion 140.80 (140.50 to 141.50): V





#13

Acrolein

Concen: 0.47 ug/l

RT: 2.29 min Scan# 280

Delta R.T. -0.01 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56


Instrument :
ClientSampleId :
VX0209WBL01

Tgt Ion: 56 Resp: 131


Ion Ratio Lower Upper

56 100



55 105.3 56.4 84.6#



Abundance Ion 56.00 (55.70 to 56.70): VX0
Ion 55.00 (54.70 to 55.70): VX0



Time-->



#15

Acrylonitrile

Concen: 0.30 ug/l

RT: 3.15 min Scan# 421

Delta R.T. -0.01 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56


Tgt Ion: 53 Resp: 282

Ion Ratio Lower Upper


53 100

52 77.0 66.5 99.7


51 28.0 29.1 43.7#




Abundance Ion 53.00 (52.70 to 53.70): VX0
Ion 52.00 (51.70 to 52.70): VX0
Ion 51.00 (50.70 to 51.70): VX0



Time-->





#16

Acetone

Concen: 0.70 ug/l

RT: 2.45 min Scan# 305

Delta R.T. -0.01 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56





Instrument :
ClientSampleId :
VX0209WBL01

Tgt Ion: 43 Resp: 571

Ion Ratio Lower Upper

43 100

58 39.4 25.9 38.9#



#17

Carbon Disulfide

Concen: 0.54 ug/l

RT: 2.57 min Scan# 326

Delta R.T. -0.00 min

Lab File: VX000050.D




Acq: 09 Feb 2018 18:56


Tgt Ion: 76 Resp: 2012

Ion Ratio Lower Upper

76 100

78 14.6 7.5 11.3#





#21

trans-1,2-Dichloroethene

Concen: 0.23 ug/l

RT: 3.17 min Scan# 423





Delta R.T. -0.01 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56

Instrument :
ClientSampled :
VX0209WBL01

Tgt Ion: 96 Resp: 321
Ion Ratio Lower Upper
96 100
61 54.5 104.9 157.3#
98 46.7 52.4 78.6#



#33

1,2-Dichloroethane-d4

Concen: 51.69 ug/l




RT: 6.08 min Scan# 901


Delta R.T. -0.01 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56

Tgt Ion: 65 Resp: 70872
Ion Ratio Lower Upper
65 100
67 51.7 0.0 106.0









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.87 min Scan# 1031
 Delta R.T. -0.00 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56




Instrument :
 ClientSampled :
 VX0209WBL01


Tgt Ion:114 Resp: 212895
 Ion Ratio Lower Upper
 114 100
 63 18.7 0.0 36.0
 88 14.7 0.0 30.4



#35
 Dibromofluoromethane
 Concen: 49.00 ug/l
 RT: 5.51 min Scan# 808
 Delta R.T. -0.01 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56

Tgt Ion:113 Resp: 60913
 Ion Ratio Lower Upper
 113 100
 111 102.2 80.5 120.7
 192 20.1 16.9 25.3









#36
 1,1-Dichloropropene
 Concen: 0.23 ug/l
 RT: 5.81 min Scan# 856
 Delta R.T. -0.01 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56




Instrument :
 ClientSampleId :
 VX0209WBL01


Tgt Ion: 75 Resp: 423
 Ion Ratio Lower Upper
 75 100
 110 0.0 19.4 58.4#
 77 13.0 24.2 36.4#



#38
 Carbon Tetrachloride
 Concen: 6.11 ug/l
 RT: 5.68 min Scan# 835
 Delta R.T. -0.12 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56

Tgt Ion: 117 Resp: 12715
 Ion Ratio Lower Upper
 117 100
 119 5.9 76.9 115.3#
 121 0.0 25.8 38.8#









#39
Methylcyclohexane
Concen: 0.20 ug/l
RT: 7.47 min Scan# 1129
Delta R.T. -0.00 min
Lab File: VX000050.D
Acq: 09 Feb 2018 18:56




Instrument :
ClientSampleId :
VX0209WBL01


Tgt Ion: 83 Resp: 452
Ion Ratio Lower Upper
83 100
55 76.4 55.7 83.5
98 43.8 39.7 59.5



#50
Toluene-d8
Concen: 47.34 ug/l
RT: 8.73 min Scan# 1335
Delta R.T. -0.00 min
Lab File: VX000050.D
Acq: 09 Feb 2018 18:56

Tgt Ion: 98 Resp: 219002
Ion Ratio Lower Upper
98 100
100 63.8 51.8 77.6





#60

Dibromochloromethane

Concen: 0.23 ug/l

RT: 9.34 min Scan# 1435

Delta R.T. -0.26 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56


Instrument :
ClientSampleId :
VX0209WBL01

Tgt Ion:129 Resp: 394




Ion Ratio Lower Upper

129 100

127 0.0 38.6 115.8#



Abundance Ion 128.80 (128.50 to 129.50): V
Ion 126.80 (126.50 to 127.50): V



#62

4-Bromofluorobenzene

Concen: 45.91 ug/l

RT: 11.15 min Scan# 1732

Delta R.T. -0.00 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56


Tgt Ion: 95 Resp: 86370

Ion Ratio Lower Upper



95 100


174 90.9 0.0 185.6

176 87.1 0.0 181.8



Abundance Ion 94.90 (94.60 to 95.60): VX0
Ion 173.80 (173.50 to 174.50): V
Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1564
Delta R.T. -0.00 min
Lab File: VX000050.D
Acq: 09 Feb 2018 18:56

Instrument :
ClientSampleId :
VX0209WBL01

Tgt Ion:117 Resp: 196998
Ion Ratio Lower Upper
117 100
82 52.4 40.3 60.5
119 32.5 25.3 37.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#64
Tetrachloroethene
Concen: 0.27 ug/l
RT: 9.34 min Scan# 1436
Delta R.T. -0.01 min
Lab File: VX000050.D
Acq: 09 Feb 2018 18:56

Tgt Ion:164 Resp: 402
Ion Ratio Lower Upper
164 100
166 110.2 103.7 155.5
129 79.2 72.2 108.4
131 96.2 69.1 103.7



Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V









#68
m/p-Xylenes
Concen: 0.29 ug/l
RT: 10.37 min Scan# 1604
Delta R.T. -0.00 min
Lab File: VX000050.D
Acq: 09 Feb 2018 18:56

Instrument :
ClientSampleId :
VX0209WBL01

Tgt Ion:106 Resp: 771
Ion Ratio Lower Upper
106 100
91 199.5 158.3 237.5




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.09 min Scan# 1887
Delta R.T. -0.00 min
Lab File: VX000050.D
Acq: 09 Feb 2018 18:56

Tgt Ion:152 Resp: 104341
Ion Ratio Lower Upper
152 100
115 57.1 39.3 117.8
150 157.8 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V









#76
 1,2,3-Trichloropropane
 Concen: 22.49 ug/l
 RT: 11.15 min Scan# 1732
 Delta R.T. -0.16 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56




Instrument :
 ClientSampled :
 VX0209WBL01


Tgt Ion: 75 Resp: 44194
 Ion Ratio Lower Upper
 75 100
 77 1.1 23.7 71.1#



#78
 n-propylbenzene
 Concen: 0.22 ug/l
 RT: 11.37 min Scan# 1769
 Delta R.T. -0.00 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56

Tgt Ion: 91 Resp: 1652
 Ion Ratio Lower Upper
 91 100
 120 21.2 12.2 36.6





#81

trans-1,4-Dichloro-2-butene

Concen: 55.65 ug/l

RT: 11.15 min Scan# 1732

Delta R.T. 0.06 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56

Instrument :

ClientSampleId :

VX0209WBL01


Tgt Ion: 75 Resp: 44194

Ion Ratio Lower Upper

75 100

53 0.0 70.3 105.5#




89 0.0 48.6 72.8#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#82

4-Chlorotoluene

Concen: 0.26 ug/l

RT: 11.52 min Scan# 1793

Delta R.T. -0.00 min

Lab File: VX000050.D


Acq: 09 Feb 2018 18:56

Tgt Ion: 91 Resp: 1377

Ion Ratio Lower Upper



91 100


126 26.5 16.3 48.8



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V









#85
 sec-Butylbenzene
 Concen: 0.20 ug/l
 RT: 11.96 min Scan# 1865
 Delta R.T. -0.00 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56

Instrument :
 ClientSampled :
 VX0209WBL01

Tgt Ion:105 Resp: 1400
 Ion Ratio Lower Upper
 105 100
 134 20.9 10.6 31.8




Abundance Ion 105.00 (104.70 to 105.70): V
 Ion 134.00 (133.70 to 134.70): V






#86
 p-Isopropyltoluene
 Concen: 0.25 ug/l
 RT: 12.07 min Scan# 1884
 Delta R.T. -0.00 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56

Tgt Ion:119 Resp: 1531
 Ion Ratio Lower Upper
 119 100
 134 33.0 14.0 41.9
 91 45.5 11.6 34.7#



Abundance Ion 119.00 (118.70 to 119.70): V
 Ion 134.00 (133.70 to 134.70): V
 Ion 91.00 (90.70 to 91.70): VX0









#87
 1,3-Dichlorobenzene
 Concen: 0.26 ug/l
 RT: 12.03 min Scan# 1877
 Delta R.T. -0.01 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56




Instrument :
 ClientSampleId :
 VX0209WBL01


Tgt Ion:146 Resp: 873
 Ion Ratio Lower Upper
 146 100
 111 49.1 18.9 56.9
 148 80.1 31.8 95.4



#88
 1,4-Dichlorobenzene
 Concen: 0.25 ug/l
 RT: 12.03 min Scan# 1877
 Delta R.T. -0.08 min
 Lab File: VX000050.D
 Acq: 09 Feb 2018 18:56

Tgt Ion:146 Resp: 873
 Ion Ratio Lower Upper
 146 100
 111 49.1 18.7 56.1
 148 80.1 32.1 96.5









#89
n-Butylbenzene
Concen: 0.37 ug/l
RT: 12.40 min Scan# 1937
Delta R.T. -0.00 min
Lab File: VX000050.D
Acq: 09 Feb 2018 18:56




Instrument :
ClientSampled :
VX0209WBL01


Tgt Ion: 91 Resp: 2118
Ion Ratio Lower Upper
91 100
92 50.6 26.1 78.2
134 25.2 14.1 42.2



#93
1,2,4-Trichlorobenzene
Concen: 0.43 ug/l
RT: 13.66 min Scan# 2144
Delta R.T. -0.00 min
Lab File: VX000050.D
Acq: 09 Feb 2018 18:56

Tgt Ion: 180 Resp: 1100
Ion Ratio Lower Upper
180 100
182 91.8 48.0 144.0
145 26.9 14.7 44.1





#94

Hexachlorobutadiene

Concen: 0.39 ug/l

RT: 13.79 min Scan# 2166

Delta R.T. -0.00 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56

Instrument :

ClientSampleId :

VX0209WBL01


Tgt Ion:225 Resp: 463

Ion Ratio Lower Upper




225 100

223 63.1 31.9 95.5

227 68.3 31.6 95.0



Abundance Ion 224.70 (224.40 to 225.40): V
Ion 222.70 (222.40 to 223.40): V
Ion 226.70 (226.40 to 227.40): V



#96

1,2,3-Trichlorobenzene

Concen: 0.27 ug/l

RT: 14.03 min Scan# 2205

Delta R.T. -0.00 min

Lab File: VX000050.D

Acq: 09 Feb 2018 18:56


Tgt Ion:180 Resp: 657

Ion Ratio Lower Upper

180 100

182 99.5 47.9 143.7

145 25.9 15.7 47.0



Abundance Ion 179.80 (179.50 to 180.50): V
Ion 181.80 (181.50 to 182.50): V
Ion 144.90 (144.60 to 145.60): V

