

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX020918\
 Data File : VX000056.D
 Acq On : 09 Feb 2018 22:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 LabSampled :
 VSTDCCC050

Quant Time: Feb 10 00:38:18 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X020918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 09 23:01:05 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	50.000	48.420	3.2	93	0.00
3 P	Chloromethane	50.000	48.966	2.1	92	0.00
4 C	Vinyl Chloride	50.000	50.092	-0.2#	93	0.00
5 T	Bromomethane	50.000	64.392	-28.8#	128	0.00
6 T	Chloroethane	50.000	52.097	-4.2	97	0.00
7 T	Trichlorofluoromethane	50.000	49.484	1.0	93	0.00
8 T	Diethyl Ether	50.000	48.866	2.3	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.020	2.0	95	0.00
10 T	Methyl Iodide	50.000	46.813	6.4	85	0.00
11 T	Tert butyl alcohol	250.000	270.919	-8.4	110	0.00
12 CM	1,1-Dichloroethene	50.000	48.395	3.2#	93	0.00
13 T	Acrolein	250.000	242.110	3.2	93	0.00
14 T	Allyl chloride	50.000	48.213	3.6	95	0.00
15 T	Acrylonitrile	250.000	259.474	-3.8	103	0.00
16 T	Acetone	250.000	271.766	-8.7	116	0.00
17 T	Carbon Disulfide	50.000	48.923	2.2	93	0.00
18 T	Methyl Acetate	50.000	54.309	-8.6	104	0.00
19 T	Methyl tert-butyl Ether	50.000	49.489	1.0	93	0.00
20 T	Methylene Chloride	50.000	49.072	1.9	94	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.827	2.3	94	0.00
22 T	Diisopropyl ether	50.000	48.552	2.9	95	0.00
23 T	Vinyl Acetate	250.000	249.805	0.1	95	0.00
24 P	1,1-Dichloroethane	50.000	48.320	3.4	92	0.00
25 T	2-Butanone	250.000	273.862	-9.5	110	0.00
26 T	2,2-Dichloropropane	50.000	46.814	6.4	89	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.127	-0.3	94	0.00
28 T	Bromochloromethane	50.000	47.750	4.5	93	0.00
29 T	Tetrahydrofuran	250.000	269.444	-7.8	108	0.00
30 C	Chloroform	50.000	50.730	-1.5#	96	0.00
31 T	Cyclohexane	50.000	49.412	1.2	95	0.00
32 T	1,1,1-Trichloroethane	50.000	50.745	-1.5	96	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.705	8.6	89	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	46.719	6.6	86	0.00
36 T	1,1-Dichloropropene	50.000	48.363	3.3	95	0.00
37 T	Ethyl Acetate	50.000	50.279	-0.6	102	0.00
38 T	Carbon Tetrachloride	50.000	49.921	0.2	95	0.00
39 T	Methylcyclohexane	50.000	48.571	2.9	94	0.00
40 TM	Benzene	50.000	49.293	1.4	94	0.00
41 T	Methacrylonitrile	50.000	50.049	-0.1	101	0.00
42 TM	1,2-Dichloroethane	50.000	49.857	0.3	96	0.00
43 T	Isopropyl Acetate	50.000	49.155	1.7	97	0.00
44 TM	Trichloroethene	50.000	49.168	1.7	94	0.00
45 C	1,2-Dichloropropane	50.000	50.324	-0.6#	95	0.00

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX020918\
 Data File : VX000056.D
 Acq On : 09 Feb 2018 22:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 LabSampled :
 VSTDCCC050

Quant Time: Feb 10 00:38:18 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X020918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 09 23:01:05 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 T	Dibromomethane	50.000	49.991	0.0	94	0.00
47 T	Bromodichloromethane	50.000	49.680	0.6	94	0.00
48 T	Methyl methacrylate	50.000	48.928	2.1	97	0.00
49 T	1,4-Dioxane	1000.000	1012.506	-1.3	100	0.00
50 S	Toluene-d8	50.000	45.477	9.0	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	262.952	-5.2	101	0.00
52 CM	Toluene	50.000	50.020	-0.0#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	49.759	0.5	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.106	-2.2	93	0.00
55 T	1,1,2-Trichloroethane	50.000	49.678	0.6	95	0.00
56 T	Ethyl methacrylate	50.000	51.116	-2.2	94	0.00
57 T	1,3-Dichloropropane	50.000	50.362	-0.7	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	253.065	-1.2	96	0.00
59 T	2-Hexanone	250.000	270.608	-8.2	103	0.00
60 T	Dibromochloromethane	50.000	51.582	-3.2	95	0.00
61 T	1,2-Dibromoethane	50.000	51.043	-2.1	94	0.00
62 S	4-Bromofluorobenzene	50.000	46.244	7.5	88	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	51.675	-3.3	98	0.00
65 PM	Chlorobenzene	50.000	49.380	1.2	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.777	-1.6	94	0.00
67 C	Ethyl Benzene	50.000	49.720	0.6#	95	0.00
68 T	m/p-Xylenes	100.000	100.436	-0.4	95	0.00
69 T	o-Xylene	50.000	50.106	-0.2	94	0.00
70 T	Styrene	50.000	49.637	0.7	93	0.00
71 P	Bromoform	50.000	51.823	-3.6	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	48.287	3.4	93	0.00
74 T	N-amyl acetate	50.000	47.635	4.7	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.945	2.1	94	0.00
76 T	1,2,3-Trichloropropane	50.000	50.952	-1.9	98	0.00
77 T	Bromobenzene	50.000	49.295	1.4	94	0.00
78 T	n-propylbenzene	50.000	48.573	2.9	95	0.00
79 T	2-Chlorotoluene	50.000	48.617	2.8	93	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.400	3.2	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.226	1.5	91	0.00
82 T	4-Chlorotoluene	50.000	48.341	3.3	94	0.00
83 T	tert-Butylbenzene	50.000	48.652	2.7	93	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.304	1.4	94	0.00
85 T	sec-Butylbenzene	50.000	48.615	2.8	94	0.00
86 T	p-Isopropyltoluene	50.000	48.503	3.0	93	0.00
87 T	1,3-Dichlorobenzene	50.000	48.080	3.8	92	0.00
88 T	1,4-Dichlorobenzene	50.000	48.339	3.3	93	0.00
89 T	n-Butylbenzene	50.000	47.482	5.0	92	0.00

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX020918\
Data File : VX000056.D
Acq On : 09 Feb 2018 22:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
LabSampleId :
VSTDCCC050

Quant Time: Feb 10 00:38:18 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X020918W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 09 23:01:05 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	50.000	48.269	3.5	92	0.00
91 T	1,2-Dichlorobenzene	50.000	48.483	3.0	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.436	3.1	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.513	5.0	92	0.00
94 T	Hexachlorobutadiene	50.000	47.502	5.0	89	0.00
95 T	Naphthalene	50.000	49.019	2.0	94	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.168	3.7	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6