

Data Path : W:\HPCHEM1\MSVOA\_X\DATA\VX021218\  
 Data File : VX000065.D  
 Acq On : 12 Feb 2018 14:52  
 Operator : JC/MD  
 Sample : MDL 03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 ClientSampled :  
 MDL 03

Quant Time: Feb 13 02:01:28 2018  
 Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X020918W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 09 23:01:05 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 5.68  | 168  | 116065   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.87  | 114  | 192414   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.12 | 117  | 181166   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.09 | 152  | 95696    | 50.00 | ug/l  | 0.00     |

|                             |        |     |          |       |         |      |
|-----------------------------|--------|-----|----------|-------|---------|------|
| System Monitoring Compounds |        |     |          |       |         |      |
| 33) 1,2-Dichloroethane-d4   | 6.09   | 65  | 62418    | 51.47 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 102.94% |      |
| 35) Dibromofluoromethane    | 5.52   | 113 | 52916    | 47.10 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 94.20%  |      |
| 50) Toluene-d8              | 8.73   | 98  | 191928   | 45.90 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 91.80%  |      |
| 62) 4-Bromofluorobenzene    | 11.15  | 95  | 77074    | 45.33 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 90.66%  |      |

| Target Compounds              |      |     |       |      |        | Qvalue |
|-------------------------------|------|-----|-------|------|--------|--------|
| 2) Dichlorodifluoromethane    | 1.20 | 85  | 864   | 0.82 | ug/l   | 94     |
| 3) Chloromethane              | 1.33 | 50  | 914   | 0.85 | ug/l # | 86     |
| 4) Vinyl Chloride             | 1.41 | 62  | 1043  | 0.86 | ug/l   | 90     |
| 5) Bromomethane               | 1.65 | 94  | 1181  | 1.83 | ug/l   | 99     |
| 6) Chloroethane               | 1.73 | 64  | 1680  | 2.32 | ug/l # | 81     |
| 7) Trichlorofluoromethane     | 1.93 | 101 | 1682  | 0.80 | ug/l   | 88     |
| 8) Diethyl Ether              | 2.19 | 74  | 646   | 0.85 | ug/l # | 46     |
| 9) 1,1,2-Trichlorotrifluoroet | 2.39 | 101 | 1010  | 0.85 | ug/l   | 96     |
| 10) Methyl Iodide             | 2.52 | 142 | 902   | 2.30 | ug/l   | 99     |
| 11) Tert butyl alcohol        | 3.07 | 59  | 2466  | 5.08 | ug/l # | 83     |
| 12) 1,1-Dichloroethene        | 2.38 | 96  | 901   | 0.80 | ug/l   | 95     |
| 13) Acrolein                  | 2.30 | 56  | 1320  | 5.30 | ug/l   | 83     |
| 14) Allyl chloride            | 2.74 | 41  | 2014  | 1.03 | ug/l   | 91     |
| 15) Acrylonitrile             | 3.16 | 53  | 3671  | 4.34 | ug/l # | 92     |
| 16) Acetone                   | 2.45 | 43  | 3763  | 5.24 | ug/l   | 92     |
| 17) Carbon Disulfide          | 2.57 | 76  | 2532  | 0.77 | ug/l # | 79     |
| 18) Methyl Acetate            | 2.79 | 43  | 1610  | 0.89 | ug/l # | 87     |
| 19) Methyl tert-butyl Ether   | 3.21 | 73  | 3161  | 0.82 | ug/l   | 98     |
| 20) Methylene Chloride        | 2.87 | 84  | 1212  | 1.01 | ug/l # | 77     |
| 21) trans-1,2-Dichloroethene  | 3.17 | 96  | 1028  | 0.85 | ug/l   | 92     |
| 22) Diisopropyl ether         | 3.88 | 45  | 2594  | 0.80 | ug/l # | 94     |
| 23) Vinyl Acetate             | 3.83 | 43  | 10812 | 3.69 | ug/l # | 94     |
| 24) 1,1-Dichloroethane        | 3.70 | 63  | 1491  | 0.74 | ug/l # | 83     |
| 25) 2-Butanone                | 4.72 | 43  | 4751  | 4.38 | ug/l   | 100    |
| 26) 2,2-Dichloropropane       | 4.60 | 77  | 1382  | 0.80 | ug/l   | 82     |
| 27) cis-1,2-Dichloroethene    | 4.61 | 96  | 984   | 0.83 | ug/l   | 98     |
| 28) Bromochloromethane        | 5.03 | 49  | 507   | 0.76 | ug/l # | 86     |
| 29) Tetrahydrofuran           | 5.20 | 42  | 3677  | 4.95 | ug/l   | 92     |
| 30) Chloroform                | 5.23 | 83  | 1633  | 0.80 | ug/l # | 79     |
| 31) Cyclohexane               | 5.59 | 56  | 1383  | 0.84 | ug/l   | 88     |
| 32) 1,1,1-Trichloroethane     | 5.51 | 97  | 1409  | 0.74 | ug/l # | 48     |
| 36) 1,1-Dichloropropene       | 5.81 | 75  | 1459  | 0.87 | ug/l   | 86     |
| 37) Ethyl Acetate             | 4.89 | 43  | 1899  | 0.88 | ug/l # | 86     |
| 38) Carbon Tetrachloride      | 5.80 | 117 | 1564  | 0.83 | ug/l   | 96     |

Data Path : W:\HPCHEM1\MSVOA\_X\DATA\VX021218\  
 Data File : VX000065.D  
 Acq On : 12 Feb 2018 14:52  
 Operator : JC/MD  
 Sample : MDL 03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 ClientSampled :  
 MDL 03

Quant Time: Feb 13 02:01:28 2018  
 Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X020918W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 09 23:01:05 2018  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc  | Units  | Dev(Min) |
|--------------------------------|-------|------|----------|-------|--------|----------|
| 39) Methylcyclohexane          | 7.47  | 83   | 1467     | 0.72  | ug/l # | 82       |
| 40) Benzene                    | 6.15  | 78   | 3647     | 0.77  | ug/l   | 92       |
| 41) Methacrylonitrile          | 5.08  | 41   | 1094     | 0.98  | ug/l # | 82       |
| 42) 1,2-Dichloroethane         | 6.23  | 62   | 1241     | 0.73  | ug/l # | 74       |
| 43) Isopropyl Acetate          | 6.48  | 43   | 2592     | 0.79  | ug/l   | 96       |
| 44) Trichloroethene            | 7.23  | 130  | 1032     | 0.70  | ug/l   | 100      |
| 45) 1,2-Dichloropropane        | 7.52  | 63   | 873      | 0.76  | ug/l # | 63       |
| 46) Dibromomethane             | 7.67  | 93   | 690      | 0.78  | ug/l   | 87       |
| 47) Bromodichloromethane       | 7.91  | 83   | 1232     | 0.71  | ug/l # | 83       |
| 48) Methyl methacrylate        | 7.79  | 41   | 1279     | 0.80  | ug/l   | 87       |
| 49) 1,4-Dioxane                | 7.77  | 88   | 532      | 12.47 | ug/l # | 74       |
| 51) 4-Methyl-2-Pentanone       | 8.66  | 43   | 8311     | 3.66  | ug/l   | 98       |
| 52) Toluene                    | 8.79  | 92   | 2413     | 0.76  | ug/l   | 95       |
| 53) t-1,3-Dichloropropene      | 8.45  | 75   | 1475     | 0.73  | ug/l # | 86       |
| 54) cis-1,3-Dichloropropene    | 9.06  | 75   | 1403     | 0.71  | ug/l   | 91       |
| 55) 1,1,2-Trichloroethane      | 9.23  | 97   | 914      | 0.69  | ug/l # | 78       |
| 56) Ethyl methacrylate         | 9.19  | 69   | 1429     | 0.70  | ug/l # | 81       |
| 57) 1,3-Dichloropropane        | 9.38  | 76   | 1484     | 0.72  | ug/l   | 95       |
| 58) 2-Chloroethyl Vinyl ether  | 8.32  | 63   | 3516     | 3.48  | ug/l   | 93       |
| 59) 2-Hexanone                 | 9.51  | 43   | 7267     | 3.93  | ug/l   | 90       |
| 60) Dibromochloromethane       | 9.59  | 129  | 935      | 0.61  | ug/l   | 94       |
| 61) 1,2-Dibromoethane          | 9.68  | 107  | 1020     | 0.69  | ug/l   | 97       |
| 64) Tetrachloroethene          | 9.34  | 164  | 986      | 0.73  | ug/l   | 84       |
| 65) Chlorobenzene              | 10.15 | 112  | 2705     | 0.72  | ug/l # | 88       |
| 66) 1,1,1,2-Tetrachloroethane  | 10.23 | 131  | 905      | 0.67  | ug/l # | 66       |
| 67) Ethyl Benzene              | 10.26 | 91   | 4614     | 0.74  | ug/l   | 100      |
| 68) m/p-Xylenes                | 10.37 | 106  | 3467     | 1.39  | ug/l   | 100      |
| 69) o-Xylene                   | 10.71 | 106  | 1657     | 0.69  | ug/l   | 93       |
| 70) Styrene                    | 10.72 | 104  | 2756     | 0.68  | ug/l   | 96       |
| 71) Bromoform                  | 10.87 | 173  | 724      | 0.57  | ug/l # | 94       |
| 73) Isopropylbenzene           | 11.03 | 105  | 4636     | 0.76  | ug/l   | 99       |
| 74) N-amyl acetate             | 6.48  | 43   | 2592     | 0.85  | ug/l # | 96       |
| 75) 1,1,2,2-Tetrachloroethane  | 11.28 | 83   | 1502     | 0.73  | ug/l   | 96       |
| 76) 1,2,3-Trichloropropane     | 11.30 | 75   | 1956     | 1.09  | ug/l   | 83       |
| 77) Bromobenzene               | 11.26 | 156  | 1255     | 0.76  | ug/l   | 91       |
| 78) n-propylbenzene            | 11.37 | 91   | 5791     | 0.82  | ug/l   | 96       |
| 79) 2-Chlorotoluene            | 11.43 | 91   | 3239     | 0.80  | ug/l   | 95       |
| 80) 1,3,5-Trimethylbenzene     | 11.52 | 105  | 3749     | 0.74  | ug/l   | 97       |
| 81) trans-1,4-Dichloro-2-buten | 11.15 | 75   | 39660    | 54.45 | ug/l # | 11       |
| 82) 4-Chlorotoluene            | 11.52 | 91   | 3750     | 0.77  | ug/l   | 97       |
| 83) tert-Butylbenzene          | 11.78 | 119  | 3782     | 0.74  | ug/l   | 90       |
| 84) 1,2,4-Trimethylbenzene     | 11.82 | 105  | 3908     | 0.75  | ug/l   | 99       |
| 85) sec-Butylbenzene           | 11.96 | 105  | 4603     | 0.73  | ug/l   | 97       |
| 86) p-Isopropyltoluene         | 12.07 | 119  | 4189     | 0.74  | ug/l   | 88       |
| 87) 1,3-Dichlorobenzene        | 12.04 | 146  | 2312     | 0.74  | ug/l   | 96       |
| 88) 1,4-Dichlorobenzene        | 12.04 | 146  | 2312     | 0.73  | ug/l   | 95       |
| 89) n-Butylbenzene             | 12.40 | 91   | 4021     | 0.76  | ug/l   | 97       |
| 90) Hexachloroethane           | 12.60 | 117  | 636      | 0.64  | ug/l   | 95       |
| 91) 1,2-Dichlorobenzene        | 12.40 | 146  | 2352     | 0.77  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropropan | 13.01 | 75   | 469      | 0.79  | ug/l   | 85       |

Data Path : W:\HPCHEM1\MSVOA\_X\DATA\VX021218\  
Data File : VX000065.D  
Acq On : 12 Feb 2018 14:52  
Operator : JC/MD  
Sample : MDL 03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
ClientSampleId :  
MDL 03

Quant Time: Feb 13 02:01:28 2018  
Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X020918W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 09 23:01:05 2018  
Response via : Initial Calibration


| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 93) 1,2,4-Trichlorobenzene | 13.66 | 180  | 1833     | 0.79 | ug/l  | 92       |
| 94) Hexachlorobutadiene    | 13.79 | 225  | 818      | 0.75 | ug/l  | 95       |
| 95) Naphthalene            | 13.84 | 128  | 5821     | 0.77 | ug/l  | 99       |
| 96) 1,2,3-Trichlorobenzene | 14.03 | 180  | 1719     | 0.76 | ug/l  | 99       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX021218\  
Data File : VX000065.D  
Acq On    : 12 Feb 2018   14:52  
Operator  : JC/MD  
Sample    : MDL 03  
Misc      : 5.0mL/MSVOA_X/WATER  
ALS Vial  : 9      Sample Multiplier: 1
```

Instrument :  
ClientSampleId :  
MDL 03

Quant Time: Feb 13 02:01:28 2018  
Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X020918W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 09 23:01:05 2018  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.68 min Scan# 835

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :


MDL 03

Tgt Ion:168 Resp: 116065

Ion Ratio Lower Upper




168 100

99 52.9 39.4 59.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#2

Dichlorodifluoromethane

Concen: 0.82 ug/l

RT: 1.20 min Scan# 100

Delta R.T. -0.00 min

Lab File: VX000065.D


Acq: 12 Feb 2018 14:52

Tgt Ion: 85 Resp: 864

Ion Ratio Lower Upper



85 100


87 28.2 15.7 47.0



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0





#3

Chloromethane

Concen: 0.85 ug/l

RT: 1.33 min Scan# 122

Delta R.T. 0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :





MDL 03

Tgt Ion: 50 Resp: 914

Ion Ratio Lower Upper

50 100

52 38.9 25.0 37.6#



#4

Vinyl Chloride

Concen: 0.86 ug/l

RT: 1.41 min Scan# 135

Delta R.T. -0.00 min

Lab File: VX000065.D




Acq: 12 Feb 2018 14:52


Tgt Ion: 62 Resp: 1043

Ion Ratio Lower Upper

62 100

64 26.0 25.4 38.0





#5

Bromomethane

Concen: 1.83 ug/l

RT: 1.65 min Scan# 174

Delta R.T. 0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :


MDL 03

Tgt Ion: 94 Resp: 1181

Ion Ratio Lower Upper




94 100

96 94.7 76.7 115.1



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: 2.32 ug/l

RT: 1.73 min Scan# 188

Delta R.T. 0.01 min

Lab File: VX000065.D


Acq: 12 Feb 2018 14:52

Tgt Ion: 64 Resp: 1680

Ion Ratio Lower Upper



64 100


66 22.2 26.4 39.6#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0









#7

Trichlorofluoromethane  
 Concen: 0.80 ug/l  
 RT: 1.93 min Scan# 221  
 Delta R.T. 0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Instrument :  
 ClientSampled :  
 MDL 03




Tgt Ion: 101 Resp: 1682  
 Ion Ratio Lower Upper  
 101 100  
 103 75.2 52.3 78.5




#8

Diethyl Ether  
 Concen: 0.85 ug/l  
 RT: 2.19 min Scan# 263  
 Delta R.T. -0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 74 Resp: 646  
 Ion Ratio Lower Upper  
 74 100  
 45 138.5 44.3 132.8#







#9

1,1,2-Trichlorotrifluoroethane

Concen: 0.85 ug/l

RT: 2.39 min Scan# 296

Delta R.T. 0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion:101 Resp: 1010

Ion Ratio Lower Upper

101 100

85 53.3 36.6 54.8




151 85.7 68.8 103.2



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 2.30 ug/l

RT: 2.52 min Scan# 317

Delta R.T. 0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52


Tgt Ion:142 Resp: 902

Ion Ratio Lower Upper

142 100

127 45.7 36.8 55.2



141 13.7 11.4 17.0




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol  
 Concen: 5.08 ug/l  
 RT: 3.07 min Scan# 408  
 Delta R.T. -0.02 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Instrument :  
 ClientSampled :  
 MDL 03




Tgt Ion: 59 Resp: 2466  
 Ion Ratio Lower Upper  
 59 100  
 57 16.8 8.4 12.6#




#12

1,1-Dichloroethene  
 Concen: 0.80 ug/l  
 RT: 2.38 min Scan# 294  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 96 Resp: 901  
 Ion Ratio Lower Upper  
 96 100  
 61 142.1 120.3 180.5  
 98 65.7 51.0 76.4









#13  
 Acrolein  
 Concen: 5.30 ug/l  
 RT: 2.30 min Scan# 281  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03


Tgt Ion: 56 Resp: 1320  
 Ion Ratio Lower Upper  
 56 100  
 55 84.3 56.4 84.6



#14  
 Allyl chloride  
 Concen: 1.03 ug/l  
 RT: 2.74 min Scan# 353  
 Delta R.T. 0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 41 Resp: 2014  
 Ion Ratio Lower Upper  
 41 100  
 39 62.1 55.8 83.6  
 76 31.0 28.7 43.1





#15

Acrylonitrile

Concen: 4.34 ug/l

RT: 3.16 min Scan# 422

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion: 53 Resp: 3671

Ion Ratio Lower Upper




53 100

52 88.1 66.5 99.7

51 43.7 29.1 43.7#



Abundance Ion 53.00 (52.70 to 53.70): VX0  
Ion 52.00 (51.70 to 52.70): VX0  
Ion 51.00 (50.70 to 51.70): VX0



#16

Acetone

Concen: 5.24 ug/l

RT: 2.45 min Scan# 306

Delta R.T. -0.01 min

Lab File: VX000065.D


Acq: 12 Feb 2018 14:52

Tgt Ion: 43 Resp: 3763



Ion Ratio Lower Upper


43 100

58 36.8 25.9 38.9



Abundance Ion 43.00 (42.70 to 43.70): VX0  
Ion 58.00 (57.70 to 58.70): VX0





#17

Carbon Disulfide

Concen: 0.77 ug/l

RT: 2.57 min Scan# 326

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :


MDL 03

Tgt Ion: 76 Resp: 2532

Ion Ratio Lower Upper

76 100

78 17.0 7.5 11.3#



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

2.57



1500

1000

500

0

Time--> 2.50 2.55 2.60 2.65



#18

Methyl Acetate

Concen: 0.89 ug/l

RT: 2.79 min Scan# 361

Delta R.T. -0.00 min

Lab File: VX000065.D


Acq: 12 Feb 2018 14:52

Tgt Ion: 43 Resp: 1610

Ion Ratio Lower Upper

43 100

74 18.9 20.2 30.4#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0

2.79

1200

1000

800


600


400

200

0

Time--> 2.75 2.80 2.85









#19

Methyl tert-butyl Ether  
Concen: 0.82 ug/l  
RT: 3.21 min Scan# 431  
Delta R.T. -0.01 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Instrument :  
ClientSampled :  
MDL 03




Tgt Ion: 73 Resp: 3161  
Ion Ratio Lower Upper  
73 100  
57 19.6 16.2 24.4




#20

Methylene Chloride  
Concen: 1.01 ug/l  
RT: 2.87 min Scan# 374  
Delta R.T. 0.01 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 84 Resp: 1212  
Ion Ratio Lower Upper  
84 100  
49 79.7 94.4 141.6#  
51 44.1 28.3 42.5#  
86 59.8 51.6 77.4





#21

trans-1,2-Dichloroethene

Concen: 0.85 ug/l

RT: 3.17 min Scan# 424

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion: 96 Resp: 1028

Ion Ratio Lower Upper

96 100

61 127.0 104.9 157.3




98 77.9 52.4 78.6



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 0.80 ug/l

RT: 3.88 min Scan# 540

Delta R.T. -0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Tgt Ion: 45 Resp: 2594


Ion Ratio Lower Upper

45 100

43 62.5 49.6 74.4

87 22.2 25.9 38.9#

59 10.9 9.2 13.8






Abundance Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 3.69 ug/l

RT: 3.83 min Scan# 532

Delta R.T. -0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :


MDL 03

Tgt Ion: 43 Resp: 10812

Ion Ratio Lower Upper




43 100

86 9.1 9.3 13.9#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0



#24

1,1-Dichloroethane

Concen: 0.74 ug/l

RT: 3.70 min Scan# 511

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52


Tgt Ion: 63 Resp: 1491

Ion Ratio Lower Upper

63 100

98 0.0 3.3 9.9#



100 8.3 2.1 6.2#




Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0







#25

2-Butanone

Concen: 4.38 ug/l

RT: 4.72 min Scan# 678

Delta R.T. -0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :





MDL 03

Tgt Ion: 43 Resp: 4751

Ion Ratio Lower Upper

43 100

72 28.6 23.0 34.6



#26

2,2-Dichloropropane

Concen: 0.80 ug/l

RT: 4.60 min Scan# 658

Delta R.T. 0.01 min

Lab File: VX000065.D




Acq: 12 Feb 2018 14:52


Tgt Ion: 77 Resp: 1382

Ion Ratio Lower Upper

77 100

97 14.8 11.8 35.4





#27

cis-1,2-Dichloroethene

Concen: 0.83 ug/l

RT: 4.61 min Scan# 660

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion: 96 Resp: 984

Ion Ratio Lower Upper

96 100

61 137.0 0.0 274.4




98 70.4 0.0 132.4



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#28

Bromochloromethane

Concen: 0.76 ug/l

RT: 5.03 min Scan# 729

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52


Tgt Ion: 49 Resp: 507

Ion Ratio Lower Upper

49 100

129 0.0 0.0 5.0



130 81.3 75.6 113.4




Abundance Ion 48.90 (48.60 to 49.60): VX0

Ion 128.80 (128.50 to 129.50): V

Ion 129.80 (129.50 to 130.50): V









#29  
Tetrahydrofuran  
Concen: 4.95 ug/l  
RT: 5.20 min Scan# 756  
Delta R.T. 0.01 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampleId :  
MDL 03


Tgt Ion: 42 Resp: 3677  
Ion Ratio Lower Upper  
42 100  
72 42.3 39.0 58.6  
71 41.9 36.7 55.1



#30  
Chloroform  
Concen: 0.80 ug/l  
RT: 5.23 min Scan# 761  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 83 Resp: 1633  
Ion Ratio Lower Upper  
83 100  
85 81.6 51.9 77.9#









#31  
Cyclohexane  
Concen: 0.84 ug/l  
RT: 5.59 min Scan# 820  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampled :  
MDL 03


Tgt Ion: 56 Resp: 1383  
Ion Ratio Lower Upper  
56 100  
69 34.3 24.5 36.7  
84 80.0 74.8 112.2



#32  
1,1,1-Trichloroethane  
Concen: 0.74 ug/l  
RT: 5.51 min Scan# 807  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 97 Resp: 1409  
Ion Ratio Lower Upper  
97 100  
99 0.0 50.8 76.2#  
61 43.6 32.1 48.1









#33  
 1,2-Dichloroethane-d4  
 Concen: 51.47 ug/l  
 RT: 6.09 min Scan# 902  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03


Tgt Ion: 65 Resp: 62418  
 Ion Ratio Lower Upper  
 65 100  
 67 51.0 0.0 106.0



#34  
 1,4-Difluorobenzene  
 Concen: 50.00 ug/l  
 RT: 6.87 min Scan# 1031  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 114 Resp: 192414  
 Ion Ratio Lower Upper  
 114 100  
 63 18.7 0.0 36.0  
 88 14.6 0.0 30.4





#35

Dibromofluoromethane

Concen: 47.10 ug/l

RT: 5.52 min Scan# 809

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion:113 Resp: 52916

Ion Ratio Lower Upper

113 100

111 100.7 80.5 120.7




192 19.5 16.9 25.3



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 0.87 ug/l

RT: 5.81 min Scan# 857

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52


Tgt Ion: 75 Resp: 1459

Ion Ratio Lower Upper

75 100

110 23.4 19.4 58.4



77 30.4 24.2 36.4




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37  
 Ethyl Acetate  
 Concen: 0.88 ug/l  
 RT: 4.89 min Scan# 705  
 Delta R.T. 0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03


Tgt Ion: 43 Resp: 1899  
 Ion Ratio Lower Upper  
 43 100  
 61 10.1 11.4 17.2#  
 70 4.1 9.0 13.6#



#38  
 Carbon Tetrachloride  
 Concen: 0.83 ug/l  
 RT: 5.80 min Scan# 855  
 Delta R.T. 0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 117 Resp: 1564  
 Ion Ratio Lower Upper  
 117 100  
 119 99.2 76.9 115.3  
 121 29.5 25.8 38.8









#39  
Methylcyclohexane  
Concen: 0.72 ug/l  
RT: 7.47 min Scan# 1129  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampled :  
MDL 03

Tgt Ion: 83 Resp: 1467  
Ion Ratio Lower Upper  
83 100  
55 86.3 55.7 83.5#  
98 40.3 39.7 59.5




#40  
Benzene  
Concen: 0.77 ug/l  
RT: 6.15 min Scan# 912  
Delta R.T. -0.01 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 78 Resp: 3647  
Ion Ratio Lower Upper  
78 100  
77 20.2 19.4 29.2











#41  
Methacrylonitrile  
Concen: 0.98 ug/l  
RT: 5.08 min Scan# 737  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampled :  
MDL 03


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 41    | Resp: | 1094  |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 39       | 53.4  | 45.8  | 68.8  |
| 67       | 57.4  | 60.6  | 91.0# |
| 52       | 12.9  | 25.4  | 38.2# |



#42  
1,2-Dichloroethane  
Concen: 0.73 ug/l  
RT: 6.23 min Scan# 925  
Delta R.T. 0.02 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 62    | Resp: | 1241  |
| Ion      | Ratio | Lower | Upper |
| 62       | 100   |       |       |
| 98       | 0.0   | 0.0   | 18.6  |





#43

Isopropyl Acetate

Concen: 0.79 ug/l

RT: 6.48 min Scan# 967

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion: 43 Resp: 2592

Ion Ratio Lower Upper

43 100

61 18.2 15.9 23.9




87 12.5 11.8 17.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#44

Trichloroethene

Concen: 0.70 ug/l

RT: 7.23 min Scan# 1089

Delta R.T. 0.01 min

Lab File: VX000065.D


Acq: 12 Feb 2018 14:52

Tgt Ion: 130 Resp: 1032

Ion Ratio Lower Upper



130 100


95 92.8 0.0 185.2



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0









#45  
 1,2-Dichloropropane  
 Concen: 0.76 ug/l  
 RT: 7.52 min Scan# 1138  
 Delta R.T. -0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03


Tgt Ion: 63 Resp: 873  
 Ion Ratio Lower Upper  
 63 100  
 65 49.8 23.7 35.5#



#46  
 Dibromomethane  
 Concen: 0.78 ug/l  
 RT: 7.67 min Scan# 1162  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 93 Resp: 690  
 Ion Ratio Lower Upper  
 93 100  
 95 84.2 66.3 99.5  
 174 90.4 90.4 135.6





#47

Bromodichloromethane

Concen: 0.71 ug/l

RT: 7.91 min Scan# 1201





Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :  
ClientSampled :  
MDL 03

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 1232  |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 85       | 80.4  | 52.3  | 78.5# |
| 127      | 8.8   | 7.4   | 11.0  |



#48

Methyl methacrylate

Concen: 0.80 ug/l




RT: 7.79 min Scan# 1181


Delta R.T. -0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 41    | Resp: | 1279  |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 69       | 77.6  | 72.1  | 108.1 |
| 39       | 71.9  | 49.6  | 74.4  |









#49  
1,4-Dioxane  
Concen: 12.47 ug/l  
RT: 7.77 min Scan# 1179  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampled :  
MDL 03


Tgt Ion: 88 Resp: 532  
Ion Ratio Lower Upper  
88 100  
43 64.5 22.6 33.8#  
58 57.1 49.4 74.2



#50  
Toluene-d8  
Concen: 45.90 ug/l  
RT: 8.73 min Scan# 1335  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 98 Resp: 191928  
Ion Ratio Lower Upper  
98 100  
100 64.6 51.8 77.6





#51

4-Methyl-2-Pentanone

Concen: 3.66 ug/l

RT: 8.66 min Scan# 1325

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52


Instrument :  
ClientSampled :  
MDL 03

Tgt Ion: 43 Resp: 8311

Ion Ratio Lower Upper




43 100

58 39.0 32.1 48.1



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#52

Toluene

Concen: 0.76 ug/l

RT: 8.79 min Scan# 1346

Delta R.T. -0.00 min

Lab File: VX000065.D


Acq: 12 Feb 2018 14:52

Tgt Ion: 92 Resp: 2413

Ion Ratio Lower Upper



92 100


91 168.4 139.8 209.6



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0









#53  
t-1,3-Dichloropropene  
Concen: 0.73 ug/l  
RT: 8.45 min Scan# 1289  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampled :  
MDL 03


Tgt Ion: 75 Resp: 1475  
Ion Ratio Lower Upper  
75 100  
77 39.0 25.0 37.4#



#54  
cis-1,3-Dichloropropene  
Concen: 0.71 ug/l  
RT: 9.06 min Scan# 1389  
Delta R.T. 0.01 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 75 Resp: 1403  
Ion Ratio Lower Upper  
75 100  
77 35.6 25.1 37.7  
39 52.1 36.1 54.1









#55  
 1,1,2-Trichloroethane  
 Concen: 0.69 ug/l  
 RT: 9.23 min Scan# 1418  
 Delta R.T. 0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 97    | Resp: | 914   |
| Ion      | Ratio | Lower | Upper |
| 97       | 100   |       |       |
| 83       | 62.2  | 66.5  | 99.7# |
| 85       | 38.7  | 43.8  | 65.6# |
| 99       | 49.1  | 52.2  | 78.4# |




#56  
 Ethyl methacrylate  
 Concen: 0.70 ug/l  
 RT: 9.19 min Scan# 1411  
 Delta R.T. -0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 69    | Resp: | 1429  |
| Ion      | Ratio | Lower | Upper |
| 69       | 100   |       |       |
| 41       | 76.3  | 49.5  | 74.3# |
| 39       | 49.3  | 29.9  | 44.9# |











#57  
 1,3-Dichloropropane  
 Concen: 0.72 ug/l  
 RT: 9.38 min Scan# 1443  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03


Tgt Ion: 76 Resp: 1484  
 Ion Ratio Lower Upper  
 76 100  
 78 35.0 25.9 38.9



#58  
 2-Chloroethyl Vinyl ether  
 Concen: 3.48 ug/l  
 RT: 8.32 min Scan# 1269  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 63 Resp: 3516  
 Ion Ratio Lower Upper  
 63 100  
 106 27.0 24.7 37.1









#59  
2-Hexanone  
Concen: 3.93 ug/l  
RT: 9.51 min Scan# 1464  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampled :  
MDL 03


Tgt Ion: 43 Resp: 7267  
Ion Ratio Lower Upper  
43 100  
58 48.0 27.6 82.7



#60  
Dibromochloromethane  
Concen: 0.61 ug/l  
RT: 9.59 min Scan# 1477  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 129 Resp: 935  
Ion Ratio Lower Upper  
129 100  
127 82.8 38.6 115.8









#61  
 1,2-Dibromoethane  
 Concen: 0.69 ug/l  
 RT: 9.68 min Scan# 1492  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03


Tgt Ion: 107 Resp: 1020  
 Ion Ratio Lower Upper  
 107 100  
 109 90.3 74.6 111.8



#62  
 4-Bromofluorobenzene  
 Concen: 45.33 ug/l  
 RT: 11.15 min Scan# 1732  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 95 Resp: 77074  
 Ion Ratio Lower Upper  
 95 100  
 174 86.3 0.0 185.6  
 176 85.1 0.0 181.8





#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 10.12 min Scan# 1564

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion:117 Resp: 181166

Ion Ratio Lower Upper

117 100

82 53.0 40.3 60.5




119 32.1 25.3 37.9



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#64

Tetrachloroethene

Concen: 0.73 ug/l

RT: 9.34 min Scan# 1436

Delta R.T. -0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Tgt Ion:164 Resp: 986


Ion Ratio Lower Upper

164 100

166 104.8 103.7 155.5

129 82.3 72.2 108.4

131 70.2 69.1 103.7






Abundance Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V









#65  
Chlorobenzene  
Concen: 0.72 ug/l  
RT: 10.15 min Scan# 1569  
Delta R.T. 0.01 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampled :  
MDL 03


Tgt Ion:112 Resp: 2705  
Ion Ratio Lower Upper  
112 100  
114 25.4 25.6 38.4#



#66  
1,1,1,2-Tetrachloroethane  
Concen: 0.67 ug/l  
RT: 10.23 min Scan# 1582  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion:131 Resp: 905  
Ion Ratio Lower Upper  
131 100  
133 93.4 47.6 142.9  
119 0.0 31.7 95.1#









#67  
Ethyl Benzene  
Concen: 0.74 ug/l  
RT: 10.26 min Scan# 1587  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Instrument :  
ClientSampled :  
MDL 03

Tgt Ion: 91 Resp: 4614  
Ion Ratio Lower Upper  
91 100  
106 32.5 25.9 38.9




Abundance Ion 91.00 (90.70 to 91.70): VX0  
Ion 106.00 (105.70 to 106.70): V






#68  
m/p-Xylenes  
Concen: 1.39 ug/l  
RT: 10.37 min Scan# 1604  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 106 Resp: 3467  
Ion Ratio Lower Upper  
106 100  
91 197.3 158.3 237.5



Abundance Ion 106.00 (105.70 to 106.70): V  
Ion 91.00 (90.70 to 91.70): VX0









#69  
o-Xylene  
Concen: 0.69 ug/l  
RT: 10.71 min Scan# 1660  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Instrument :  
ClientSampled :  
MDL 03

Tgt Ion:106 Resp: 1657  
Ion Ratio Lower Upper  
106 100  
91 218.3 104.0 312.0




Abundance Ion 106.00 (105.70 to 106.70): V  
Ion 91.00 (90.70 to 91.70): VX0






#70  
Styrene  
Concen: 0.68 ug/l  
RT: 10.72 min Scan# 1662  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion:104 Resp: 2756  
Ion Ratio Lower Upper  
104 100  
78 49.8 37.7 56.5  
103 57.9 44.0 66.0



Abundance Ion 104.00 (103.70 to 104.70): V  
Ion 78.00 (77.70 to 78.70): VX0  
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 0.57 ug/l

RT: 10.87 min Scan# 1686

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion:173 Resp: 724

Ion Ratio Lower Upper

173 100

175 51.8 23.9 71.8




254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.09 min Scan# 1887

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52


Tgt Ion:152 Resp: 95696

Ion Ratio Lower Upper

152 100

115 58.1 39.3 117.8



150 157.5 0.0 349.2




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V







#73

Isopropylbenzene

Concen: 0.76 ug/l

RT: 11.03 min Scan# 1713

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :


MDL 03

Tgt Ion: 105 Resp: 4636

Ion Ratio Lower Upper




105 100

120 26.7 13.8 41.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 0.85 ug/l

RT: 6.48 min Scan# 967

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Tgt Ion: 43 Resp: 2592


Ion Ratio Lower Upper

43 100

70 0.0 0.0 0.0

55 0.0 0.1 0.1#

61 18.2 15.9 23.9






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0





#75

1,1,2,2-Tetrachloroethane

Concen: 0.73 ug/l

RT: 11.28 min Scan# 1754

Delta R.T. -0.00 min

Lab File: VX000065.D





Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 1502  |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 131      | 14.8  | 5.5   | 16.7  |
| 85       | 67.6  | 32.6  | 97.7  |



#76

1,2,3-Trichloropropane

Concen: 1.09 ug/l




RT: 11.30 min Scan# 1758


Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 75    | Resp: | 1956  |
| Ion      | Ratio | Lower | Upper |
| 75       | 100   |       |       |
| 77       | 35.9  | 23.7  | 71.1  |





#77

Bromobenzene

Concen: 0.76 ug/l

RT: 11.26 min Scan# 1751

Delta R.T. -0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03


Tgt Ion:156 Resp: 1255

Ion Ratio Lower Upper

156 100

77 148.3 66.2 198.6




158 95.5 48.6 145.9



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 0.82 ug/l

RT: 11.37 min Scan# 1768

Delta R.T. -0.01 min

Lab File: VX000065.D


Acq: 12 Feb 2018 14:52

Tgt Ion: 91 Resp: 5791

Ion Ratio Lower Upper



91 100


120 22.7 12.2 36.6



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 0.80 ug/l

RT: 11.43 min Scan# 1778

Delta R.T. -0.01 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :


MDL 03

Tgt Ion: 91 Resp: 3239

Ion Ratio Lower Upper




91 100

126 33.3 18.2 54.6



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V



#80

1,3,5-Trimethylbenzene

Concen: 0.74 ug/l

RT: 11.52 min Scan# 1793

Delta R.T. -0.00 min

Lab File: VX000065.D


Acq: 12 Feb 2018 14:52

Tgt Ion: 105 Resp: 3749

Ion Ratio Lower Upper



105 100


120 51.7 24.8 74.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V









#81  
trans-1,4-Dichloro-2-butene  
Concen: 54.45 ug/l  
RT: 11.15 min Scan# 1732  
Delta R.T. 0.06 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52




Instrument :  
ClientSampleId :  
MDL 03


Tgt Ion: 75 Resp: 39660  
Ion Ratio Lower Upper  
75 100  
53 0.0 70.3 105.5#  
89 0.0 48.6 72.8#



#82  
4-Chlorotoluene  
Concen: 0.77 ug/l  
RT: 11.52 min Scan# 1794  
Delta R.T. 0.01 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion: 91 Resp: 3750  
Ion Ratio Lower Upper  
91 100  
126 30.8 16.3 48.8









#83  
 tert-Butylbenzene  
 Concen: 0.74 ug/l  
 RT: 11.78 min Scan# 1836  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Instrument :  
 ClientSampled :  
 MDL 03

Tgt Ion:119 Resp: 3782  
 Ion Ratio Lower Upper  
 119 100  
 91 64.4 27.9 83.7  
 134 26.0 11.7 35.1




Abundance Ion 119.00 (118.70 to 119.70): V  
 Ion 91.00 (90.70 to 91.70): V  
 Ion 134.00 (133.70 to 134.70): V






#84  
 1,2,4-Trimethylbenzene  
 Concen: 0.75 ug/l  
 RT: 11.82 min Scan# 1842  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion:105 Resp: 3908  
 Ion Ratio Lower Upper  
 105 100  
 120 46.3 23.4 70.2



Abundance Ion 105.00 (104.70 to 105.70): V  
 Ion 120.00 (119.70 to 120.70): V









#85  
 sec-Butylbenzene  
 Concen: 0.73 ug/l  
 RT: 11.96 min Scan# 1865  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03


Tgt Ion:105 Resp: 4603  
 Ion Ratio Lower Upper  
 105 100  
 134 22.8 10.6 31.8



#86  
 p-Isopropyltoluene  
 Concen: 0.74 ug/l  
 RT: 12.07 min Scan# 1884  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion:119 Resp: 4189  
 Ion Ratio Lower Upper  
 119 100  
 134 30.8 14.0 41.9  
 91 32.3 11.6 34.7









#87  
 1,3-Dichlorobenzene  
 Concen: 0.74 ug/l  
 RT: 12.04 min Scan# 1878  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03

Tgt Ion:146 Resp: 2312  
 Ion Ratio Lower Upper  
 146 100  
 111 43.9 18.9 56.9  
 148 63.2 31.8 95.4




#88  
 1,4-Dichlorobenzene  
 Concen: 0.73 ug/l  
 RT: 12.04 min Scan# 1878  
 Delta R.T. -0.07 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion:146 Resp: 2312  
 Ion Ratio Lower Upper  
 146 100  
 111 43.9 18.7 56.1  
 148 63.2 32.1 96.5







#89

n-Butylbenzene

Concen: 0.76 ug/l

RT: 12.40 min Scan# 1937

Delta R.T. -0.00 min

Lab File: VX000065.D

Acq: 12 Feb 2018 14:52

Instrument :

ClientSampled :

MDL 03





Tgt Ion: 91 Resp: 4021

Ion Ratio Lower Upper

91 100

92 50.0 26.1 78.2

134 26.6 14.1 42.2



#90

Hexachloroethane

Concen: 0.64 ug/l

RT: 12.60 min Scan# 1971

Delta R.T. -0.00 min

Lab File: VX000065.D




Acq: 12 Feb 2018 14:52


Tgt Ion: 117 Resp: 636

Ion Ratio Lower Upper

117 100

201 82.9 43.5 130.7









#91  
 1,2-Dichlorobenzene  
 Concen: 0.77 ug/l  
 RT: 12.40 min Scan# 1938  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52




Instrument :  
 ClientSampled :  
 MDL 03


Tgt Ion:146 Resp: 2352  
 Ion Ratio Lower Upper  
 146 100  
 111 37.1 19.4 58.2  
 148 64.8 32.0 96.2



#92  
 1,2-Dibromo-3-Chloropropane  
 Concen: 0.79 ug/l  
 RT: 13.01 min Scan# 2038  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion: 75 Resp: 469  
 Ion Ratio Lower Upper  
 75 100  
 155 81.4 52.0 156.0  
 157 118.8 65.3 195.9









#93  
 1,2,4-Trichlorobenzene  
 Concen: 0.79 ug/l  
 RT: 13.66 min Scan# 2144  
 Delta R.T. -0.00 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Instrument :  
 ClientSampled :  
 MDL 03

Tgt Ion:180 Resp: 1833  
 Ion Ratio Lower Upper  
 180 100  
 182 88.1 48.0 144.0  
 145 32.6 14.7 44.1




Abundance Ion 179.80 (179.50 to 180.50): V  
 Ion 181.80 (181.50 to 182.50): V  
 Ion 144.80 (144.50 to 145.50): V






#94  
 Hexachlorobutadiene  
 Concen: 0.75 ug/l  
 RT: 13.79 min Scan# 2165  
 Delta R.T. -0.01 min  
 Lab File: VX000065.D  
 Acq: 12 Feb 2018 14:52

Tgt Ion:225 Resp: 818  
 Ion Ratio Lower Upper  
 225 100  
 223 59.8 31.9 95.5  
 227 59.8 31.6 95.0



Abundance Ion 224.70 (224.40 to 225.40): V  
 Ion 222.70 (222.40 to 223.40): V  
 Ion 226.70 (226.40 to 227.40): V









#95  
Naphthalene  
Concen: 0.77 ug/l  
RT: 13.84 min Scan# 2174  
Delta R.T. -0.01 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Instrument :  
ClientSampled :  
MDL 03

Tgt Ion:128 Resp: 5821  
Ion Ratio Lower Upper  
128 100  
127 13.2 10.3 15.5  
129 10.7 8.6 13.0



#96  
1,2,3-Trichlorobenzene  
Concen: 0.76 ug/l  
RT: 14.03 min Scan# 2205  
Delta R.T. -0.00 min  
Lab File: VX000065.D  
Acq: 12 Feb 2018 14:52

Tgt Ion:180 Resp: 1719  
Ion Ratio Lower Upper  
180 100  
182 96.7 47.9 143.7  
145 31.9 15.7 47.0

