


Data Path : W:\HPCHEM1\MSVOA X\Data\VX021318\
Data File : VX000072.D
Acq On : 13 Feb 2018 09:09
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ClientSampled :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X020918W.M
Title : SW846 8260
Last Update : Fri Feb 09 23:01:05 2018



AutoFind: Scans 1731, 1732, 1733; Background Corrected with Scan 1725

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	3738	PASS
75	95	30	60	51.3	10730	PASS
95	95	100	100	100.0	20933	PASS
96	95	5	9	7.2	1499	PASS
173	174	0.00	2	1.3	244	PASS
174	95	50	100	86.9	18193	PASS
175	174	5	9	7.7	1392	PASS
176	174	95	101	96.9	17621	PASS
177	176	5	9	6.9	1211	PASS