

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022118\
 Data File : VX000121.D
 Acq On : 21 Feb 2018 17:09
 Operator : JC/MD
 Sample : VX0221WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ClientSampleId :
 VX0221WBL01

Quant Time: Feb 22 07:13:24 2018
 Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X022018W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Feb 21 06:17:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	14436	30.00	ug/l	-0.01
28) 1,4-Difluorobenzene	6.87	114	89742	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	85774	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	35190	31.50	ug/l	-0.01
Spiked Amount	30.000	Range	50 - 169	Recovery	=	105.00%
60) 4-Bromofluorobenzene	11.15	95	39141	27.89	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.97%
63) Toluene-d8	8.72	98	110854	30.17	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.57%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022118\
Data File : VX000121.D
Acq On : 21 Feb 2018 17:09
Operator : JC/MD
Sample : VX0221WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ClientSampleId :
VX0221WBL01

Quant Time: Feb 22 07:13:24 2018
Quant Method : W:\HPCHEM1\MSVOA X\METHOD\624X022018W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Feb 21 06:17:26 2018
Response via : Initial Calibration

