


Data Path : W:\HPCHEM1\MSVOA X\Data\VX022218\
Data File : VX000127.D
Acq On : 22 Feb 2018 11:59
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\624X022018W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Feb 21 06:17:26 2018



AutoFind: Scans 1731, 1732, 1733; Background Corrected with Scan 1725

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	4496	PASS
75	95	30	60	55.0	12198	PASS
95	95	100	100	100.0	22181	PASS
96	95	5	9	7.2	1604	PASS
173	174	0.00	2	1.3	246	PASS
174	95	50	100	87.1	19320	PASS
175	174	5	9	7.5	1450	PASS
176	174	95	101	96.1	18565	PASS
177	176	5	9	6.7	1238	PASS