

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022718\
 Data File : VX000143.D
 Acq On : 27 Feb 2018 14:52
 Operator : JC/MD
 Sample : VX0227WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ClientSampleId :
 VX0227WBL01

Quant Time: Feb 28 05:37:48 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 12:46:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	133125	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	209394	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	187911	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	93415	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	71395	42.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.90%	
35) Dibromofluoromethane	5.51	113	62585	47.36	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	94.72%	
50) Toluene-d8	8.73	98	248548	47.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.36%	
62) 4-Bromofluorobenzene	11.15	95	87835	42.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.94%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022718\
Data File : VX000143.D
Acq On : 27 Feb 2018 14:52
Operator : JC/MD
Sample : VX0227WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ClientSampleId :
VX0227WBL01

Quant Time: Feb 28 05:37:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 12:46:41 2018
Response via : Initial Calibration

