

Data Path : W:\HPCHEM1\MSVOA X\DATA\WX022718\
 Data File : VX000153.D
 Acq On : 27 Feb 2018 19:17
 Operator : JC/MD
 Sample : J1683-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ClientSampled :
 RFK-RMAB-MW08-GW

Manual Integrations
 APPROVED

apatel
 2/28/2018 3:54:56 PM

Quant Time: Feb 28 05:57:54 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 12:46:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	91620	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	149599	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	140420	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	98224	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	57805	50.52	ug/l	0.00
Spiked Amount						
			Recovery	=	101.04%	
35) Dibromofluoromethane	5.52	113	44731	47.38	ug/l	0.00
Spiked Amount						
			Recovery	=	94.76%	
50) Toluene-d8	8.73	98	180430	48.44	ug/l	0.00
Spiked Amount						
			Recovery	=	96.88%	
62) 4-Bromofluorobenzene	11.15	95	76915	52.67	ug/l	0.00
Spiked Amount						
			Recovery	=	105.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.08	59	8521	19.33	ug/l	# 72
16) Acetone	2.45	43	5354	7.22	ug/l	95
19) Methyl tert-butyl Ether	3.21	73	19898	6.04	ug/l	# 72
31) Cyclohexane	5.59	56	38247	26.75	ug/l	94
39) Methylcyclohexane	7.46	83	45549	28.18	ug/l	96
40) Benzene	6.15	78	101942	26.58	ug/l	99
52) Toluene	8.79	92	54047	21.09	ug/l	98
67) Ethyl Benzene	10.26	91	2681432	538.82	ug/l	98
68) m/p-Xylenes	10.37	106	3213103	1630.53	ug/l	97
69) o-Xylene	10.71	106	3070443	1615.56	ug/l	98
73) Isopropylbenzene	11.02	105	214741	34.50	ug/l	98
78) n-propylbenzene	11.37	91	577998	78.75	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	1102112	209.41	ug/l	100
84) 1,2,4-Trimethylbenzene	11.82	105	6300614	1153.03	ug/l	97
85) sec-Butylbenzene	11.96	105	41789m	6.33	ug/l	
95) Naphthalene	13.85	128	3046694	375.97	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022718\
Data File : VX000153.D
Acq On : 27 Feb 2018 19:17
Operator : JC/MD
Sample : J1683-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ClientSampleId :
RFK-RMAB-MW08-GW

Quant Time: Feb 28 05:57:54 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 12:46:41 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

apatel
2/28/2018 3:54:56 PM

