

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022818\
 Data File : VX000165.D
 Acq On : 28 Feb 2018 14:29
 Operator : JC/MD
 Sample : J1674-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-2

Quant Time: Mar 01 06:50:48 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 12:46:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	97159	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	154788	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	143034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	75303	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	60287	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
35) Dibromofluoromethane	5.51	113	46185	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
50) Toluene-d8	8.73	98	188819	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	11.15	95	66827	44.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.46%	
Target Compounds						
16) Acetone	2.45	43	2203	2.80	ug/l	# 79
19) Methyl tert-butyl Ether	3.22	73	57908	16.58	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022818\
Data File : VX000165.D
Acq On : 28 Feb 2018 14:29
Operator : JC/MD
Sample : J1674-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-2

Quant Time: Mar 01 06:50:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 12:46:41 2018
Response via : Initial Calibration

