

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022818\
 Data File : VX000168.D
 Acq On : 28 Feb 2018 15:46
 Operator : JC/MD
 Sample : J1674-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AFTER-CARBON-2

Quant Time: Mar 01 07:03:18 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 12:46:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	93937	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	150204	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	137457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	71374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	59256	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
35) Dibromofluoromethane	5.52	113	45242	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
50) Toluene-d8	8.73	98	179658	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
62) 4-Bromofluorobenzene	11.15	95	62016	42.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.58%	
Target Compounds						
16) Acetone	2.45	43	423921	557.20	ug/l	98
25) 2-Butanone	4.73	43	48298	45.47	ug/l	96
29) Tetrahydrofuran	5.18	42	622476	899.99	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022818\
Data File : VX000168.D
Acq On : 28 Feb 2018 15:46
Operator : JC/MD
Sample : J1674-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AFTER-CARBON-2

Quant Time: Mar 01 07:03:18 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 12:46:41 2018
Response via : Initial Calibration

