

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022818\
 Data File : VX000169.D
 Acq On : 28 Feb 2018 16:11
 Operator : JC/MD
 Sample : J1674-05RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FINAL-EFFLUENT

Quant Time: Mar 01 07:04:46 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 12:46:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	86909	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	140286	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	125275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	64915	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	54243	49.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.96%	
35) Dibromofluoromethane	5.52	113	40626	45.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.76%	
50) Toluene-d8	8.73	98	168256	48.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.34%	
62) 4-Bromofluorobenzene	11.15	95	59249	43.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.52%	

Target Compounds						Qvalue
16) Acetone	2.45	43	4265	6.06	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX022818\
Data File : VX000169.D
Acq On : 28 Feb 2018 16:11
Operator : JC/MD
Sample : J1674-05RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FINAL-EFFLUENT

Quant Time: Mar 01 07:04:46 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 12:46:41 2018
Response via : Initial Calibration

