

Data File : VX000200.D
Acq On : 02 Mar 2018 18:59
Operator : JC/MD
Sample : J1779-03 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-5

Quant Time: Mar 03 05:39:35 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 12:46:41 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	80684	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	131365	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	120403	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	64050	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	52428	52.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.06%	
35) Dibromofluoromethane	5.52	113	39622	47.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.58%	
50) Toluene-d8	8.73	98	161457	49.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.74%	
62) 4-Bromofluorobenzene	11.15	95	55671	43.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.82%	

Target Compounds					Qvalue	
52) Toluene	8.79	92	3928	1.75	ug/l	88
67) Ethyl Benzene	10.26	91	1996	0.47	ug/l #	77
68) m/p-Xylenes	10.37	106	3770	2.23	ug/l	95
69) o-Xylene	10.71	106	3499	2.15	ug/l	89
95) Naphthalene	13.84	128	263844	49.93	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000200.D
Acq On : 02 Mar 2018 18:59
Operator : JC/MD
Sample : J1779-03 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-5

Quant Time: Mar 03 05:39:35 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X021418W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 12:46:41 2018
Response via : Initial Calibration

