


Data Path : W:\HPCHEM1\MSVOA X\Data\VX030718\
Data File : VX000205.D
Acq On : 07 Mar 2018 11:19
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Title : SW846 8260
Last Update : Thu Mar 08 02:15:47 2018



AutoFind: Scans 1731, 1732, 1733; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	4117	PASS
75	95	30	60	55.0	11480	PASS
95	95	100	100	100.0	20861	PASS
96	95	5	9	7.8	1620	PASS
173	174	0.00	2	1.7	307	PASS
174	95	50	100	87.8	18321	PASS
175	174	5	9	7.9	1455	PASS
176	174	95	101	97.9	17941	PASS
177	176	5	9	6.5	1166	PASS