

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX030818\
 Data File : VX000232.D
 Acq On : 08 Mar 2018 20:36
 Operator : JC/MD
 Sample : J1819-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-9

Quant Time: Mar 09 06:12:21 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 08 02:15:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	4144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	6069	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	5551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	2830	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	2001	34.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.66%	
35) Dibromofluoromethane	5.53	113	1311	34.39	ug/l	0.02
Spiked Amount 50.000			Recovery	=	68.78%	
50) Toluene-d8	8.73	98	5337	33.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.60%	
62) 4-Bromofluorobenzene	11.14	95	2351	38.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.30%	
Target Compounds						
16) Acetone	2.45	43	3000	72.72	ug/l	95
20) Methylene Chloride	2.87	84	13732	279.11	ug/l	98
68) m/p-Xylenes	10.37	106	1161	15.42	ug/l	93
84) 1,2,4-Trimethylbenzene	11.81	105	1925	12.65	ug/l	97
95) Naphthalene	13.84	128	13823	63.15	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX030818\
Data File : VX000232.D
Acq On : 08 Mar 2018 20:36
Operator : JC/MD
Sample : J1819-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-9

Quant Time: Mar 09 06:12:21 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

