

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX030818\
 Data File : VX000234.D
 Acq On : 08 Mar 2018 21:26
 Operator : JC/MD
 Sample : J1819-06 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-11

Quant Time: Mar 09 06:17:34 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 08 02:15:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	4216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	6748	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	6012	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	3314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	1853	31.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.42%	
35) Dibromofluoromethane	5.50	113	1486	35.06	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	70.12%	
50) Toluene-d8	8.73	98	6385	36.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.74%	
62) 4-Bromofluorobenzene	11.14	95	2066	30.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.30%	
Target Compounds						
16) Acetone	2.45	43	3099	73.84	ug/l	# 88
95) Naphthalene	13.84	128	21795	85.03	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX030818\
Data File : VX000234.D
Acq On : 08 Mar 2018 21:26
Operator : JC/MD
Sample : J1819-06 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-11

Quant Time: Mar 09 06:17:34 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

