

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX030818\
 Data File : VX000237.D
 Acq On : 08 Mar 2018 22:43
 Operator : JC/MD
 Sample : J1823-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MHB-WCB

Quant Time: Mar 09 06:21:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 08 02:15:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	2816	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	4286	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	4212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	2477	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.09	65	2841	72.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	145.56%	
35) Dibromofluoromethane	5.52	113	1672	62.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	124.20%	
50) Toluene-d8	8.73	98	6945	62.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	124.56%	
62) 4-Bromofluorobenzene	11.14	95	2619	60.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.36%	
Target Compounds						
16) Acetone	2.45	43	3744	133.55	ug/l	96
20) Methylene Chloride	2.87	84	314019	9392.41	ug/l	99
95) Naphthalene	13.85	128	19017	99.26	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX030818\
Data File : VX000237.D
Acq On : 08 Mar 2018 22:43
Operator : JC/MD
Sample : J1823-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MHB-WCB

Quant Time: Mar 09 06:21:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

