

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX030818\
 Data File : VX000243.D
 Acq On : 09 Mar 2018 01:15
 Operator : JC/MD
 Sample : J1760-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-04-030618

Quant Time: Mar 09 05:45:38 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 08 02:15:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	46327	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	78749	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	76887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	38919	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	40721	63.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.82%	
35) Dibromofluoromethane	5.51	113	30154	60.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.92%	
50) Toluene-d8	8.73	98	116850	57.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.06%	
62) 4-Bromofluorobenzene	11.15	95	40504	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
Target Compounds						
16) Acetone	2.45	43	3409	7.39	ug/l	91
20) Methylene Chloride	2.87	84	5674	10.32	ug/l	95
95) Naphthalene	13.85	128	5465	1.82	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX030818\
Data File : VX000243.D
Acq On : 09 Mar 2018 01:15
Operator : JC/MD
Sample : J1760-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-04-030618

Quant Time: Mar 09 05:45:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

