


Data Path : W:\HPCHEM1\MSVOA X\Data\VX030818\
Data File : VX000222.D
Acq On : 08 Mar 2018 14:38
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Title : SW846 8260
Last Update : Thu Mar 08 02:15:47 2018



AutoFind: Scans 1731, 1732, 1733; Background Corrected with Scan 1725

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	2845	PASS
75	95	30	60	57.2	7546	PASS
95	95	100	100	100.0	13203	PASS
96	95	5	9	6.7	890	PASS
173	174	0.00	2	1.4	161	PASS
174	95	50	100	88.4	11669	PASS
175	174	5	9	7.4	864	PASS
176	174	95	101	100.7	11747	PASS
177	176	5	9	5.7	668	PASS