

Data File : VX000258.D
Acq On : 09 Mar 2018 16:25
Operator : JC/MD
Sample : J1839-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Mar 10 00:46:30 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	112566	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	189337	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	182685	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	95549	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.09	65	82239	52.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.40%	
35) Dibromofluoromethane	5.52	113	59816	50.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.58%	
50) Toluene-d8	8.73	98	235694	47.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.70%	
62) 4-Bromofluorobenzene	11.15	95	84085	43.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.46%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000258.D

Acq On : 09 Mar 2018 16:25

Operator : JC/MD

Sample : J1839-01

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 1 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

MW-1

Quant Time: Mar 10 00:46:30 2018

Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M

Quant Title : SW846 8260

QLast Update : Thu Mar 08 02:15:47 2018

Response via : Initial Calibration

