

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031218\
 Data File : VX000284.D
 Acq On : 12 Mar 2018 19:02
 Operator : JC/MD
 Sample : J1699-74
 Misc : 5.82g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B27-A

Quant Time: Mar 13 05:24:30 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 08 02:15:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	180738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	306002	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.13	117	293049	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	163730	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	118313	47.22	ug/l	0.00
Spiked Amount			Recovery	=	94.44%	
35) Dibromofluoromethane	5.52	113	91114	47.40	ug/l	0.00
Spiked Amount			Recovery	=	94.80%	
50) Toluene-d8	8.73	98	366100	45.99	ug/l	0.00
Spiked Amount			Recovery	=	91.98%	
62) 4-Bromofluorobenzene	11.15	95	138175	44.47	ug/l	0.00
Spiked Amount			Recovery	=	88.94%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
39) Methylcyclohexane	7.46	83	3563	1.07	ug/l	87
67) Ethyl Benzene	10.26	91	3989	0.39	ug/l	89
68) m/p-Xylenes	10.37	106	4631	1.16	ug/l	99
69) o-Xylene	10.71	106	884	0.23	ug/l	89
78) n-propylbenzene	11.37	91	4984	0.41	ug/l	99
80) 1,3,5-Trimethylbenzene	11.52	105	5205	0.61	ug/l	96
84) 1,2,4-Trimethylbenzene	11.82	105	13305	1.51	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031218\
Data File : VX000284.D
Acq On : 12 Mar 2018 19:02
Operator : JC/MD
Sample : J1699-74
Misc : 5.82g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B27-A

Quant Time: Mar 13 05:24:30 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

