

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031318\
 Data File : VX000301.D
 Acq On : 13 Mar 2018 15:47
 Operator : JC/MD
 Sample : J1898-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-11

Quant Time: Mar 14 01:11:35 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 08 02:15:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	116872	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	196978	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	185791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	99209	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	71442	44.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.20%	
35) Dibromofluoromethane	5.52	113	52961	42.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.60%	
50) Toluene-d8	8.73	98	210416	41.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.12%	
62) 4-Bromofluorobenzene	11.15	95	75377	37.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	75.36%	
Target Compounds						
16) Acetone	2.46	43	3130	2.69	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031318\
Data File : VX000301.D
Acq On : 13 Mar 2018 15:47
Operator : JC/MD
Sample : J1898-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-11

Quant Time: Mar 14 01:11:35 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

