

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031318\
 Data File : VX000306.D
 Acq On : 13 Mar 2018 17:54
 Operator : JC/MD
 Sample : J1898-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-10-COMP

Quant Time: Mar 14 01:21:40 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 08 02:15:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	115235	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	195870	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	186018	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	98275	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	81278	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
35) Dibromofluoromethane	5.52	113	59301	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
50) Toluene-d8	8.73	98	231387	45.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.82%	
62) 4-Bromofluorobenzene	11.15	95	83754	42.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.22%	
Target Compounds						
16) Acetone	2.46	43	3467	3.02	ug/l	94
20) Methylene Chloride	2.87	84	85934	62.81	ug/l	94
95) Naphthalene	13.84	128	16288	2.14	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031318\
Data File : VX000306.D
Acq On : 13 Mar 2018 17:54
Operator : JC/MD
Sample : J1898-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-10-COMP

Quant Time: Mar 14 01:21:40 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

