

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031318\
 Data File : VX000308.D
 Acq On : 13 Mar 2018 18:46
 Operator : JC/MD
 Sample : J1885-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT-3867

Quant Time: Mar 14 01:25:32 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 08 02:15:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	110999	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	190062	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	182278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	95393	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	77889	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
35) Dibromofluoromethane	5.51	113	58707	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
50) Toluene-d8	8.73	98	230250	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
62) 4-Bromofluorobenzene	11.15	95	82731	42.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.74%	
Target Compounds						
16) Acetone	2.45	43	3410	3.09	ug/l	99
20) Methylene Chloride	2.87	84	1908	1.45	ug/l #	89
95) Naphthalene	13.85	128	21643	2.93	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031318\
Data File : VX000308.D
Acq On : 13 Mar 2018 18:46
Operator : JC/MD
Sample : J1885-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT-3867

Quant Time: Mar 14 01:25:32 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 08 02:15:47 2018
Response via : Initial Calibration

