


Data Path : W:\HPCHEM1\MSVOA X\Data\VX031318\
Data File : VX000288.D
Acq On : 13 Mar 2018 09:23
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X030718W.M
Title : SW846 8260
Last Update : Thu Mar 08 02:15:47 2018



AutoFind: Scans 1731, 1732, 1733; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	3849	PASS
75	95	30	60	56.1	10906	PASS
95	95	100	100	100.0	19445	PASS
96	95	5	9	7.4	1441	PASS
173	174	0.00	2	1.1	190	PASS
174	95	50	100	85.7	16665	PASS
175	174	5	9	8.3	1379	PASS
176	174	95	101	99.6	16592	PASS
177	176	5	9	6.6	1100	PASS