

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031518\
 Data File : VX000325.D
 Acq On : 15 Mar 2018 13:26
 Operator : JC/MD
 Sample : VX0315WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0315WBL01

Quant Time: Mar 16 02:52:13 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 15 03:39:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	116525	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	195969	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	188851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	100177	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	76995	47.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.54%	
35) Dibromofluoromethane	5.51	113	60112	48.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.24%	
50) Toluene-d8	8.72	98	242121	46.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.88%	
62) 4-Bromofluorobenzene	11.15	95	86352	41.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031518\
Data File : VX000325.D
Acq On : 15 Mar 2018 13:26
Operator : JC/MD
Sample : VX0315WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0315WBL01

Quant Time: Mar 16 02:52:13 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

