

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031518\
 Data File : VX000329.D
 Acq On : 15 Mar 2018 15:28
 Operator : JC/MD
 Sample : J1808-01DL 50X
 Misc : 5.32g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-04-12-13DL

Quant Time: Mar 16 03:06:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 15 03:39:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	98629	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	165153	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	159890	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	84952	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	64975	47.62	ug/l	0.00
Spiked Amount			Recovery	=	95.24%	
35) Dibromofluoromethane	5.51	113	50744	48.70	ug/l	0.00
Spiked Amount			Recovery	=	97.40%	
50) Toluene-d8	8.73	98	207718	47.79	ug/l	0.00
Spiked Amount			Recovery	=	95.58%	
62) 4-Bromofluorobenzene	11.15	95	73900	42.39	ug/l	0.00
Spiked Amount			Recovery	=	84.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Cyclohexane	5.60	56	4640	3.01	ug/l	# 76
39) Methylcyclohexane	7.48	83	15279	8.73	ug/l	98
40) Benzene	6.17	78	1162	0.27	ug/l	# 88
52) Toluene	8.79	92	3620	1.21	ug/l	100
67) Ethyl Benzene	10.26	91	42240	7.61	ug/l	98
68) m/p-Xylenes	10.37	106	44203	20.44	ug/l	99
69) o-Xylene	10.71	106	10765	5.12	ug/l	95
73) Isopropylbenzene	11.02	105	8185	1.60	ug/l	97
78) n-propylbenzene	11.37	91	36171	5.86	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	45907	10.53	ug/l	97
84) 1,2,4-Trimethylbenzene	11.82	105	119528	26.67	ug/l	100
85) sec-Butylbenzene	11.95	105	3850	0.72	ug/l	# 56
95) Naphthalene	13.84	128	24938	3.91	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031518\
Data File : VX000329.D
Acq On : 15 Mar 2018 15:28
Operator : JC/MD
Sample : J1808-01DL 50X
Misc : 5.32g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-04-12-13DL

Quant Time: Mar 16 03:06:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

