

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031518\
 Data File : VX000333.D
 Acq On : 15 Mar 2018 17:10
 Operator : JC/MD
 Sample : PB107372TB 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB107372TB

Quant Time: Mar 16 04:12:37 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 15 03:39:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	171295	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	283122	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	271519	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	133081	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	115068	48.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.12%	
35) Dibromofluoromethane	5.51	113	88927	49.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.56%	
50) Toluene-d8	8.73	98	361242	48.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.96%	
62) 4-Bromofluorobenzene	11.15	95	121229	40.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.14%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031518\
Data File : VX000333.D
Acq On : 15 Mar 2018 17:10
Operator : JC/MD
Sample : PB107372TB 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB107372TB

Quant Time: Mar 16 04:12:37 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

