

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031518\
 Data File : VX000334.D
 Acq On : 15 Mar 2018 17:36
 Operator : JC/MD
 Sample : J1752-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NB-08-031418

Quant Time: Mar 16 04:14:15 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 15 03:39:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	169672	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	278605	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	259239	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	118591	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	110946	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
35) Dibromofluoromethane	5.51	113	87591	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
50) Toluene-d8	8.73	98	349084	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
62) 4-Bromofluorobenzene	11.15	95	113363	38.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.10%	
Target Compounds						
16) Acetone	2.45	43	12186	7.78	ug/l	97
20) Methylene Chloride	2.87	84	13434	6.62	ug/l	97
95) Naphthalene	13.84	128	89938	10.10	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX031518\
Data File : VX000334.D
Acq On : 15 Mar 2018 17:36
Operator : JC/MD
Sample : J1752-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-08-031418

Quant Time: Mar 16 04:14:15 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

