

Data Path : W:\HPCHEM1\MSVOA X\Data\VX031518\
 Data File : VX000341.D
 Acq On : 15 Mar 2018 20:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 16 04:32:53 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 15 03:39:21 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.539	0.479	11.1	94	0.00
3 P	Chloromethane	0.549	0.517	5.8	102	0.00
4 C	Vinyl Chloride	0.615	0.623	-1.3#	101	0.00
5 T	Bromomethane	0.686	0.851	-24.1#	124	0.00
6 T	Chloroethane	0.437	0.416	4.8	105	0.00
7 T	Trichlorofluoromethane	1.068	1.101	-3.1	105	0.00
8 T	Diethyl Ether	0.380	0.389	-2.4	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.575	0.589	-2.4	104	0.00
10 T	Methyl Iodide	0.581	0.557	4.1	90	0.00
11 T	Tert butyl alcohol	0.224	0.245	-9.4	115	0.00
12 CM	1,1-Dichloroethene	0.546	0.549	-0.5#	103	0.00
13 T	Acrolein	0.096	0.113	-17.7	127	0.00
14 T	Allyl chloride	0.973	0.998	-2.6	106	0.00
15 T	Acrylonitrile	0.432	0.468	-8.3	111	0.00
16 T	Acetone	0.462	0.502	-8.7	109	0.00
17 T	Carbon Disulfide	1.594	1.545	3.1	99	0.00
18 T	Methyl Acetate	1.034	1.137	-10.0	114	0.00
19 T	Methyl tert-butyl Ether	1.921	2.023	-5.3	108	0.00
20 T	Methylene Chloride	0.598	0.625	-4.5	107	0.00
21 T	trans-1,2					

Data Path : W:\HPCHEM1\MSVOA X\Data\VX031518\
 Data File : VX000341.D
 Acq On : 15 Mar 2018 20:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 16 04:32:53 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 15 03:39:21 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.255	0.259	-1.6	106	0.00
47 T	Bromodichloromethane	0.453	0.491	-8.4	109	0.00
48 T	Methyl methacrylate	0.441	0.463	-5.0	109	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	108	0.00
50 S	Toluene-d8	1.316	1.245	5.4	94	0.00
51 T	4-Methyl-2-Pentanone	0.677	0.749	-10.6	113	0.00
52 CM	Toluene	0.905	0.923	-2.0#	106	0.00
53 T	t-1,3-Dichloropropene	0.523	0.563	-7.6	105	0.00
54 T	cis-1,3-Dichloropropene	0.512	0.549	-7.2	102	0.00
55 T	1,1,2-Trichloroethane	0.363	0.375	-3.3	106	0.00
56 T	Ethyl methacrylate	0.538	0.594	-10.4	109	0.00
57 T	1,3-Dichloropropane	0.589	0.609	-3.4	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.247	0.275	-11.3	108	0.00
59 T	2-Hexanone	0.570	0.640	-12.3	113	0.00
60 T	Dibromochloromethane	0.384	0.417	-8.6	106	0.00
61 T	1,2-Dibromoethane	0.403	0.422	-4.7	105	0.00
62 S	4-Bromofluorobenzene	0.528	0.491	7.0	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.357	0.347	2.8	104	0.00
65 PM	Chlorobenzene	1.018	1.050	-		

Data Path : W:\HPCHEM1\MSVOA X\Data\VX031518\
Data File : VX000341.D
Acq On : 15 Mar 2018 20:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 16 04:32:53 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.447	0.474	-6.0	106	0.00
91 T	1,2-Dichlorobenzene	1.502	1.579	-5.1	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.290	0.320	-10.3	112	0.00
93 T	1,2,4-Trichlorobenzene	1.082	1.028	5.0	100	0.00
94 T	Hexachlorobutadiene	0.427	0.418	2.1	108	0.00
95 T	Naphthalene	3.756	3.991	-6.3	108	0.00
96 T	1,2,3-Trichlorobenzene	1.093	1.057	3.3	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6