


Data Path : W:\HPCHEM1\MSVOA X\Data\VX031518\
Data File : VX000322.D
Acq On : 15 Mar 2018 09:32
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Title : SW846 8260
Last Update : Thu Mar 15 03:39:21 2018



AutoFind: Scans 1731, 1732, 1733; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.8	4743	PASS
75	95	30	60	55.5	11570	PASS
95	95	100	100	100.0	20833	PASS
96	95	5	9	7.4	1551	PASS
173	174	0.00	2	1.4	264	PASS
174	95	50	100	90.7	18892	PASS
175	174	5	9	8.6	1627	PASS
176	174	95	101	99.0	18704	PASS
177	176	5	9	6.6	1235	PASS