

Data File : VX000344.D
Acq On : 16 Mar 2018 13:16
Operator : JC/MD
Sample : VX0316MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0316MBL01

Quant Time: Mar 16 14:50:39 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	98296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	163928	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	153726	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	70916	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	66548	48.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.88%	
35) Dibromofluoromethane	5.51	113	50500	48.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.66%	
50) Toluene-d8	8.73	98	201403	46.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.36%	
62) 4-Bromofluorobenzene	11.15	95	66237	38.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000344.D
Acq On : 16 Mar 2018 13:16
Operator : JC/MD
Sample : VX0316MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0316MBL01

Quant Time: Mar 16 14:50:39 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

