

Data File : VX000350.D
Acq On : 16 Mar 2018 16:26
Operator : JC/MD
Sample : J1808-01DL 20X
Misc : 5.32g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-04-12-13DL

Manual Integrations
APPROVED

apatel
3/19/2018 5:32:16 PM

Quant Time: Mar 17 09:53:33 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	160434	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	276398	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	271178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	164759	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	108766	49.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.02%	
35) Dibromofluoromethane	5.51	113	83434	47.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.68%	
50) Toluene-d8	8.72	98	348493	47.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.80%	
62) 4-Bromofluorobenzene	11.15	95	132576	45.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.88%	

Target Compounds						Qvalue
31) Cyclohexane	5.58	56	13787	5.50	ug/l	# 92
39) Methylcyclohexane	7.46	83	60475	20.65	ug/l	96
52) Toluene	8.79	92	21020	4.20	ug/l	95
67) Ethyl Benzene	10.26	91	291741	30.99	ug/l	97
68) m/p-Xylenes	10.37	106	335004	91.33	ug/l	98
69) o-Xylene	10.71	106	82223	23.07	ug/l	99
73) Isopropylbenzene	11.02	105	60530	6.10	ug/l	99
78) n-propylbenzene	11.37	91	280238	23.42	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	390718	46.22	ug/l	97
84) 1,2,4-Trimethylbenzene	11.82	105	1049727	120.78	ug/l	98
85) sec-Butylbenzene	11.96	105	28461m	2.76	ug/l	
86) p-Isopropyltoluene	12.07	119	22446	2.44	ug/l	99
95) Naphthalene	13.84	128	220386	17.81	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000350.D
Acq On : 16 Mar 2018 16:26
Operator : JC/MD
Sample : J1808-01DL 20X
Misc : 5.32g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
SB-04-12-13DL

Manual Integrations
APPROVED

apatel
3/19/2018 5:32:16 PM

Quant Time: Mar 17 09:53:33 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X031418W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 15 03:39:21 2018
Response via : Initial Calibration

