

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032018\
 Data File : VX000365.D
 Acq On : 20 Mar 2018 17:37
 Operator : JC/MD
 Sample : J1975-07 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FS-SOIL

Quant Time: Mar 21 05:04:12 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 20 14:01:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	99593	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	172402	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	158088	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	68677	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	70575	54.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.58%	
35) Dibromofluoromethane	5.51	113	54083	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	8.73	98	211538	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
62) 4-Bromofluorobenzene	11.15	95	66401	36.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.58%	
Target Compounds						
20) Methylene Chloride	2.86	84	15315	12.26	ug/l	99
95) Naphthalene	13.84	128	9849	1.76	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032018\
Data File : VX000365.D
Acq On : 20 Mar 2018 17:37
Operator : JC/MD
Sample : J1975-07 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FS-SOIL

Quant Time: Mar 21 05:04:12 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 20 14:01:51 2018
Response via : Initial Calibration

