

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032018\
 Data File : VX000367.D
 Acq On : 20 Mar 2018 18:28
 Operator : JC/MD
 Sample : J1760-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-03-031918

Quant Time: Mar 21 05:09:21 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 20 14:01:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	101573	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	172002	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	163476	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	76294	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	68924	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
35) Dibromofluoromethane	5.51	113	54039	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	8.73	98	212479	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
62) 4-Bromofluorobenzene	11.15	95	70770	38.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.76%	
Target Compounds						
95) Naphthalene	13.84	128	31567	5.06	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032018\
Data File : VX000367.D
Acq On : 20 Mar 2018 18:28
Operator : JC/MD
Sample : J1760-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-03-031918

Quant Time: Mar 21 05:09:21 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 20 14:01:51 2018
Response via : Initial Calibration

