

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032018\
 Data File : VX000370.D
 Acq On : 20 Mar 2018 19:44
 Operator : JC/MD
 Sample : J1979-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-1

Quant Time: Mar 21 05:16:26 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 20 14:01:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	102753	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	176392	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	168385	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	80690	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	71331	53.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.36%	
35) Dibromofluoromethane	5.52	113	55128	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	8.73	98	219433	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
62) 4-Bromofluorobenzene	11.15	95	73075	39.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.22%	
Target Compounds						
16) Acetone	2.45	43	1790	1.60	ug/l	95
20) Methylene Chloride	2.87	84	18517	14.37	ug/l	98
95) Naphthalene	13.84	128	58220	8.83	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032018\
Data File : VX000370.D
Acq On : 20 Mar 2018 19:44
Operator : JC/MD
Sample : J1979-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-1

Quant Time: Mar 21 05:16:26 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 20 14:01:51 2018
Response via : Initial Calibration

