

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032118\
 Data File : VX000375.D
 Acq On : 21 Mar 2018 10:50
 Operator : JC/MD
 Sample : VX0321WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0321WBL01

Quant Time: Mar 22 05:58:28 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	107044	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	191234	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	181431	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	86142	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	75169	53.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.60%	
35) Dibromofluoromethane	5.51	113	58344	48.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.84%	
50) Toluene-d8	8.72	98	235853	47.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.52%	
62) 4-Bromofluorobenzene	11.15	95	80064	39.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	78.90%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032118\
Data File : VX000375.D
Acq On : 21 Mar 2018 10:50
Operator : JC/MD
Sample : VX0321WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0321WBL01

Quant Time: Mar 22 05:58:28 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

