


Data Path : W:\HPCHEM1\MSVOA X\Data\VX032118\
Data File : VX000373.D
Acq On : 21 Mar 2018 09:26
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Title : SW846 8260
Last Update : Wed Mar 21 04:52:42 2018



AutoFind: Scans 1731, 1732, 1733; Background Corrected with Scan 1725

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	4875	PASS
75	95	30	60	57.6	12490	PASS
95	95	100	100	100.0	21674	PASS
96	95	5	9	7.0	1510	PASS
173	174	0.00	2	1.5	287	PASS
174	95	50	100	88.5	19186	PASS
175	174	5	9	7.3	1401	PASS
176	174	95	101	97.2	18649	PASS
177	176	5	9	7.2	1337	PASS