

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
 Data File : VX000401.D
 Acq On : 22 Mar 2018 13:06
 Operator : JC/MD
 Sample : J1992-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRUM-1-4-B

Quant Time: Mar 23 06:00:40 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	96317	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	162183	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	154314	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	72554	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	67516	53.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.40%	
35) Dibromofluoromethane	5.51	113	52089	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
50) Toluene-d8	8.73	98	204332	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
62) 4-Bromofluorobenzene	11.15	95	66706	38.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.74%	
Target Compounds						
16) Acetone	2.45	43	26321	25.17	ug/l	96
20) Methylene Chloride	2.87	84	22647	18.75	ug/l	90
25) 2-Butanone	4.72	43	4334	3.69	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
Data File : VX000401.D
Acq On : 22 Mar 2018 13:06
Operator : JC/MD
Sample : J1992-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-1-4-B

Quant Time: Mar 23 06:00:40 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

