

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
 Data File : VX000413.D
 Acq On : 22 Mar 2018 18:11
 Operator : JC/MD
 Sample : J2005-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7-031918

Quant Time: Mar 23 06:54:30 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	95764	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	165976	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	157696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	70637	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	68334	54.66	ug/l	0.00
Spiked Amount			Recovery	=	109.32%	
35) Dibromofluoromethane	5.51	113	53278	50.95	ug/l	0.00
Spiked Amount			Recovery	=	101.90%	
50) Toluene-d8	8.73	98	204543	47.22	ug/l	0.00
Spiked Amount			Recovery	=	94.44%	
62) 4-Bromofluorobenzene	11.15	95	66717	38.13	ug/l	0.00
Spiked Amount			Recovery	=	76.26%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	3171	3.05	ug/l	89
19) Methyl tert-butyl Ether	3.22	73	7989	2.08	ug/l	95
40) Benzene	6.15	78	6321	1.40	ug/l #	67
67) Ethyl Benzene	10.26	91	8834	1.48	ug/l	95
73) Isopropylbenzene	11.02	105	31483	6.73	ug/l	99
78) n-propylbenzene	11.37	91	17642	3.16	ug/l	99
84) 1,2,4-Trimethylbenzene	11.82	105	1451	0.35	ug/l	100
85) sec-Butylbenzene	11.96	105	8924	1.82	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
Data File : VX000413.D
Acq On : 22 Mar 2018 18:11
Operator : JC/MD
Sample : J2005-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7-031918

Quant Time: Mar 23 06:54:30 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

