

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
 Data File : VX000414.D
 Acq On : 22 Mar 2018 18:37
 Operator : JC/MD
 Sample : J2005-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8S-031918

Quant Time: Mar 23 06:56:06 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	86322	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	155141	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	144165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	67155	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	63844	56.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.32%	
35) Dibromofluoromethane	5.52	113	48523	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
50) Toluene-d8	8.73	98	188180	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
62) 4-Bromofluorobenzene	11.15	95	59950	36.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.78%	
Target Compounds						
11) Tert butyl alcohol	3.07	59	20948	54.05	ug/l	97
19) Methyl tert-butyl Ether	3.21	73	203502	58.81	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
Data File : VX000414.D
Acq On : 22 Mar 2018 18:37
Operator : JC/MD
Sample : J2005-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8S-031918

Quant Time: Mar 23 06:56:06 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

