

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
 Data File : VX000418.D
 Acq On : 22 Mar 2018 20:19
 Operator : JC/MD
 Sample : J2005-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1-031918

Quant Time: Mar 23 07:05:48 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	88247	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	154254	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	144381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	66089	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	63640	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
35) Dibromofluoromethane	5.51	113	49948	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
50) Toluene-d8	8.73	98	186549	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
62) 4-Bromofluorobenzene	11.15	95	59143	36.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.30%	
Target Compounds						
11) Tert butyl alcohol	3.07	59	32701	82.53	ug/l	99
16) Acetone	2.45	43	2290	2.39	ug/l #	87
19) Methyl tert-butyl Ether	3.21	73	210980	59.64	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
Data File : VX000418.D
Acq On : 22 Mar 2018 20:19
Operator : JC/MD
Sample : J2005-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1-031918

Quant Time: Mar 23 07:05:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

