

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
 Data File : VX000419.D
 Acq On : 22 Mar 2018 20:44
 Operator : JC/MD
 Sample : J1992-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRUM-1-4-A

Quant Time: Mar 23 07:08:38 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	86861	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	152839	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	145800	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	69777	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	63658	56.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.28%	
35) Dibromofluoromethane	5.51	113	49649	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
50) Toluene-d8	8.73	98	191162	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
62) 4-Bromofluorobenzene	11.15	95	61294	38.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.10%	
Target Compounds						
16) Acetone	2.45	43	26332	27.92	ug/l	Qvalue 100
25) 2-Butanone	4.73	43	4371	4.12	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032218\
Data File : VX000419.D
Acq On : 22 Mar 2018 20:44
Operator : JC/MD
Sample : J1992-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-1-4-A

Quant Time: Mar 23 07:08:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

