

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
 Data File : VX034686.D
 Acq On : 23 Mar 2023 09:51
 Operator : JC/MD
 Sample : VX0323WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0323WBL01

Quant Time: Mar 23 13:28:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	286224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	479346	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	431718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	195957	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	205546	49.363	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.720%
35) Dibromofluoromethane	5.385	113	152028	48.036	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.080%
50) Toluene-d8	8.647	98	574231	48.603	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.200%
62) 4-Bromofluorobenzene	11.079	95	222741	46.675	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034686.D
Acq On : 23 Mar 2023 09:51
Operator : JC/MD
Sample : VX0323WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0323WBL01

Quant Time: Mar 23 13:28:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

