

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
 Data File : VX034693.D
 Acq On : 23 Mar 2023 13:10
 Operator : JC/MD
 Sample : PB151618TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151618TB

Quant Time: Mar 24 05:43:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	314851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	523595	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	463384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	197300	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	223565	48.809	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.620%
35) Dibromofluoromethane	5.379	113	162478	46.999	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.000%
50) Toluene-d8	8.647	98	629858	48.806	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.079	95	235745	45.225	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.440%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	28175	13.058	ug/l	96
20) Methylene Chloride	2.788	84	13081	3.124	ug/l	98
37) Ethyl Acetate	4.708	43	67572	11.649	ug/l	98
43) Isopropyl Acetate	6.330	43	280633	28.558	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034693.D
Acq On : 23 Mar 2023 13:10
Operator : JC/MD
Sample : PB151618TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151618TB

Quant Time: Mar 24 05:43:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

