

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
 Data File : VX034694.D
 Acq On : 23 Mar 2023 13:33
 Operator : JC/MD
 Sample : 02021-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-6E-032023

Quant Time: Mar 24 05:43:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	329762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	553421	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	496272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	223590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	234461	48.873	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.740%
35) Dibromofluoromethane	5.385	113	173710	47.540	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.080%
50) Toluene-d8	8.647	98	660951	48.455	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.920%
62) 4-Bromofluorobenzene	11.079	95	257028	46.650	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.300%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	62341	27.587	ug/l	96
20) Methylene Chloride	2.788	84	23784	5.423	ug/l	97
37) Ethyl Acetate	4.714	43	66762	10.889	ug/l	97
43) Isopropyl Acetate	6.336	43	287551	27.685	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034694.D
Acq On : 23 Mar 2023 13:33
Operator : JC/MD
Sample : 02021-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-6E-032023

Quant Time: Mar 24 05:43:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

