

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
 Data File : VX034697.D
 Acq On : 23 Mar 2023 14:43
 Operator : JC/MD
 Sample : 02023-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-13-TEST-PIT-ON-GRAVELLY

Quant Time: Mar 24 05:44:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	305797	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	518378	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	452467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	199226	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	220967	49.670	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.340%
35) Dibromofluoromethane	5.385	113	162775	47.559	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.120%
50) Toluene-d8	8.647	98	610464	47.779	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.560%
62) 4-Bromofluorobenzene	11.079	95	230983	44.757	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.520%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	28554	13.626	ug/l	95
20) Methylene Chloride	2.788	84	13385	3.291	ug/l	95
37) Ethyl Acetate	4.721	43	52875	9.207	ug/l	98
43) Isopropyl Acetate	6.342	43	222126	22.831	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034697.D
Acq On : 23 Mar 2023 14:43
Operator : JC/MD
Sample : 02023-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-13-TEST-PIT-ON-GRAVELLY

Quant Time: Mar 24 05:44:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

