

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
 Data File : VX034698.D
 Acq On : 23 Mar 2023 15:06
 Operator : JC/MD
 Sample : 02023-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-12-TEST-PIT-ON-OCEAN

Quant Time: Mar 24 05:44:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	311161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	527073	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	466497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	210115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	227019	50.150	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	100.300%	
35) Dibromofluoromethane	5.385	113	166736	47.913	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.820%	
50) Toluene-d8	8.647	98	634482	48.840	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	97.680%	
62) 4-Bromofluorobenzene	11.079	95	241747	46.070	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	92.140%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	29377	13.777	ug/l	93
20) Methylene Chloride	2.788	84	13419	3.243	ug/l	95
37) Ethyl Acetate	4.715	43	62286	10.667	ug/l	98
43) Isopropyl Acetate	6.342	43	252612	25.537	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034698.D
Acq On : 23 Mar 2023 15:06
Operator : JC/MD
Sample : 02023-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-12-TEST-PIT-ON-OCEAN

Quant Time: Mar 24 05:44:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

