

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
 Data File : VX034699.D
 Acq On : 23 Mar 2023 15:29
 Operator : JC/MD
 Sample : 02045-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC1-1-20230322

Quant Time: Mar 24 05:45:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	298034	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	502173	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	439457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190174	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	213706	49.289	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.580%		
35) Dibromofluoromethane	5.385	113	158991	47.953	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.900%		
50) Toluene-d8	8.647	98	597522	48.276	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.560%		
62) 4-Bromofluorobenzene	11.079	95	218529	43.710	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.420%		
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	33251	16.281	ug/l	98
20) Methylene Chloride	2.794	84	12555	3.167	ug/l	91
37) Ethyl Acetate	4.714	43	55038	9.893	ug/l	99
40) Benzene	6.037	78	65566	4.401	ug/l	99
43) Isopropyl Acetate	6.336	43	243588	25.845	ug/l	99
52) Toluene	8.720	92	35576	3.857	ug/l	97
67) Ethyl Benzene	10.195	91	48674	2.820	ug/l	99
68) m/p-Xylenes	10.299	106	18617	2.808	ug/l	97
69) o-Xylene	10.646	106	14939	2.245	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	27803	2.207	ug/l	98
95) Naphthalene	13.774	128	146286	10.447	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034699.D
Acq On : 23 Mar 2023 15:29
Operator : JC/MD
Sample : 02045-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC1-1-20230322

Quant Time: Mar 24 05:45:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

